

PEER-REVIEW REPORT

Name of journal: *World Journal of Clinical Cases*

Manuscript NO: 69708

Title: Flip-over of blood vessel intima caused by vascular closure device: A case report

Provenance and peer review: Unsolicited manuscript; Externally peer reviewed

Peer-review model: Single blind

Reviewer's code: 05386080

Position: Peer Reviewer

Academic degree: FACP, MBBS, MD

Professional title: Assistant Professor

Reviewer's Country/Territory: United States

Author's Country/Territory: China

Manuscript submission date: 2021-07-09

Reviewer chosen by: AI Technique

Reviewer accepted review: 2021-07-23 03:01

Reviewer performed review: 2021-08-02 22:45

Review time: 10 Days and 19 Hours

Scientific quality	<input type="checkbox"/> Grade A: Excellent <input type="checkbox"/> Grade B: Very good <input checked="" type="checkbox"/> Grade C: Good <input type="checkbox"/> Grade D: Fair <input type="checkbox"/> Grade E: Do not publish
Language quality	<input checked="" type="checkbox"/> Grade A: Priority publishing <input type="checkbox"/> Grade B: Minor language polishing <input type="checkbox"/> Grade C: A great deal of language polishing <input type="checkbox"/> Grade D: Rejection
Conclusion	<input type="checkbox"/> Accept (High priority) <input type="checkbox"/> Accept (General priority) <input checked="" type="checkbox"/> Minor revision <input type="checkbox"/> Major revision <input type="checkbox"/> Rejection
Re-review	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
Peer-reviewer	Peer-Review: <input checked="" type="checkbox"/> Anonymous <input type="checkbox"/> Onymous

statements

Conflicts-of-Interest: [] Yes [Y] No

SPECIFIC COMMENTS TO AUTHORS

I appreciate the author's efforts in writing the exciting case report. Figures were of good quality, and legends were well explained. Consider including other articles that talked about other complications secondary to starClose and VCD. Another case series mentioned the thrombus secondary to the starClose device. DOI: <https://doi.org/10.3174/ajnr.A2001> It is prudent to include other complications secondary to VCDs <https://doi.org/10.1053/j.tvir.2015.04.008> this article talked about other complications secondary to VCDs. Rodriguez et al also reported pseudoaneurysm as a complication. Authors talked about more and less common complications secondary to the starClose device, and readers would also get benefit if authors include other complications secondary to the starClose device. If authors include a table containing complications secondary to the starClose device would increase the quality of the manuscript. Thank you for the opportunity.