
World Journal of Experimental Medicine

World J Exp Med 2022 September 20; 12(5): 92-107



Contents

Bimonthly Volume 12 Number 5 September 20, 2022

MINIREVIEWS

92 Bariatric surgery outcomes following organ transplantation: A review study Kheirvari M, Goudarzi H, Hemmatizadeh M, Anbara T

LETTER TO THE EDITOR

100 Performance of a serological IgM and IgG qualitative test for COVID-19 diagnosis: An experimental study

Freire de Melo F, Martins Oliveira Diniz L, Nélio Januário J, Fernando Gonçalves Ferreira J, Dórea RSDM, de Brito BB, Marques HS, Lemos FFB, Silva Luz M, Rocha Pinheiro SL, de Magalhães Queiroz DM

Diet and nutrition against inflammatory bowel disease: Trick or treat(ment)? 104

Greco S, Bonsi B, Fabbri N



Contents

Bimonthly Volume 12 Number 5 September 20, 2022

ABOUT COVER

Peer Reviewer of World Journal of Experimental Medicine, Cory J Xian, PhD, Research Professor, UniSA Clinical and Health Sciences, University of South Australia, Adelaide, SA 5001, Australia. cory.xian@unisa.edu.au

AIMS AND SCOPE

The primary aim of the World Journal of Experimental Medicine (WJEM, World J Exp Med) is to provide scholars and readers from various fields of experimental medicine with a platform to publish high-quality basic and clinical research articles and communicate their research findings online.

WJEM mainly publishes articles reporting research results and findings obtained in the field of experimental medicine and covering a wide range of topics including clinical laboratory medicine (applied and basic research in hematology, body fluid examination, cytomorphology, genetic diagnosis of hematological disorders, thrombosis and hemostasis, and blood typing and transfusion), biochemical examination (applied and basic research in laboratory automation and information system, biochemical methodology, and biochemical diagnostics), etc.

INDEXING/ABSTRACTING

The WJEM is now abstracted and indexed in PubMed, PubMed Central, Scopus, Reference Citation Analysis, China National Knowledge Infrastructure, China Science and Technology Journal Database, and Superstar Journals Database.

RESPONSIBLE EDITORS FOR THIS ISSUE

Production Editor: Hua-Ge Yu; Production Department Director: Xu Guo; Editorial Office Director: Ji-Hong Liu.

NAME OF JOURNAL

World Journal of Experimental Medicine

ISSN 2220-315x (online)

I ALINCH DATE

December 20, 2011

FREQUENCY

Bimonthly

EDITORS-IN-CHIEF

Leonardo Roever

EDITORIAL BOARD MEMBERS

https://www.wjgnet.com/2220-315x/editorialboard.htm

PUBLICATION DATE

September 20, 2022

COPYRIGHT

© 2022 Baishideng Publishing Group Inc

INSTRUCTIONS TO AUTHORS

https://www.wjgnet.com/bpg/gerinfo/204

GUIDELINES FOR ETHICS DOCUMENTS

https://www.wignet.com/bpg/GerInfo/287

GUIDELINES FOR NON-NATIVE SPEAKERS OF ENGLISH

https://www.wjgnet.com/bpg/gerinfo/240

PUBLICATION ETHICS

https://www.wignet.com/bpg/GerInfo/288

PUBLICATION MISCONDUCT

https://www.wjgnet.com/bpg/gerinfo/208

ARTICLE PROCESSING CHARGE

https://www.wignet.com/bpg/gerinfo/242

STEPS FOR SUBMITTING MANUSCRIPTS

https://www.wjgnet.com/bpg/GerInfo/239

ONLINE SUBMISSION

https://www.f6publishing.com

© 2022 Baishideng Publishing Group Inc. All rights reserved. 7041 Koll Center Parkway, Suite 160, Pleasanton, CA 94566, USA E-mail: bpgoffice@wjgnet.com https://www.wjgnet.com

Submit a Manuscript: https://www.f6publishing.com

World J Exp Med 2022 September 20; 12(5): 100-103

DOI: 10.5493/wjem.v12.i5.100

LETTER TO THE EDITOR

ISSN 2220-315x (online)

Performance of a serological IgM and IgG qualitative test for COVID-19 diagnosis: An experimental study in Brazil

Fabrício Freire de Melo, Lilian Martins Oliveira Diniz, José Nélio Januário, João Fernando Gonçalves Ferreira, Rafael Santos Dantas Miranda Dórea, Breno Bittencourt de Brito, Hanna Santos Marques, Fabian Fellipe Bueno Lemos, Marcel Silva Luz, Samuel Luca Rocha Pinheiro, Dulciene Maria de Magalhães Queiroz

Specialty type: Medicine, research and experimental

Provenance and peer review:

Unsolicited article; Externally peer

Peer-review model: Single blind

Peer-review report's scientific quality classification

Grade A (Excellent): 0 Grade B (Very good): B Grade C (Good): C, C Grade D (Fair): 0 Grade E (Poor): 0

P-Reviewer: Ghartey-Kwansah G, Ghana: Nazari N. Iran

Received: April 5, 2022

Peer-review started: April 5, 2022 First decision: June 16, 2022 **Revised:** June 27, 2022 Accepted: August 12, 2022 Article in press: August 12, 2022 Published online: September 20,

2022



Fabrício Freire de Melo, Breno Bittencourt de Brito, Fabian Fellipe Bueno Lemos, Marcel Silva Luz, Samuel Luca Rocha Pinheiro, Instituto Multidisciplinar em Saúde, Universidade Federal da Bahia, Vitória da Conquista 45029-094, Bahia, Brazil

Lilian Martins Oliveira Diniz, Departamento de Pediatria, Universidade Federal de Minas Gerais, Belo Horizonte 31270-901, Brazil

José Nélio Januário, Núcleo de Ações e Pesquisa em Apoio Diagnóstico, Universidade Federal de Minas Gerais, Belo Horizonte 31270-901, Brazil

João Fernando Gonçalves Ferreira, Laboratory of Research in Bacteriology, Universidade Federal de Minas Gerais, Belo Horizonte 31270-901, Brazil

Rafael Santos Dantas Miranda Dórea, Dulciene Maria de Magalhães Queiroz, Laboratory of Bacteriology, Faculdade de Medicina, Universidade Federal de Minas Gerais, Belo Horizonte 31270-901, Brazil

Hanna Santos Marques, Campus Vitória da Conquista, Universidade Estadual do Sudoeste da Bahia, Vitória da Conquista 45029-094, Brazil

Corresponding author: Fabrício Freire de Melo, PhD, Professor, Instituto Multidisciplinar em Saúde, Universidade Federal da Bahia, Rua Hormindo Barros, 58-Candeias, Vitória da Conquista 45029-094, Bahia, Brazil. freiremeloufba@gmail.com

Abstract

Qualitative antibody tests are an easy, point-of-care diagnostic method that is useful in diagnosing coronavirus disease 2019, especially in situations where reverse transcription-polymerase chain reaction is negative. However, some factors are able to affect its sensitivity and accuracy, which may contribute to these tests not being used as a first-line diagnostic tool.

Key Words: Serological test; IgM; IgG; COVID-19; Diagnosis; Antibody

100

©The Author(s) 2022. Published by Baishideng Publishing Group Inc. All rights reserved.



WJEM | https://www.wjgnet.com

Core Tip: In this study we compared a quantitative enzyme-linked immunosorbent assay test that detects antibodies against the severe acute respiratory syndrome coronavirus 2 S1 epitope with the qualitative test. Our results demonstrate that the quantitative tests have significantly higher sensitivity rates, evidencing limitations in the use of the qualitative antibody detection test as a first-line diagnostic tool.

Citation: Freire de Melo F, Martins Oliveira Diniz L, Nélio Januário J, Fernando Gonçalves Ferreira J, Dórea RSDM, de Brito BB, Marques HS, Lemos FFB, Silva Luz M, Rocha Pinheiro SL, de Magalhães Queiroz DM. Performance of a serological IgM and IgG qualitative test for COVID-19 diagnosis: An experimental study in Brazil. World J Exp Med 2022; 12(5): 100-103

URL: https://www.wjgnet.com/2220-315x/full/v12/i5/100.htm

DOI: https://dx.doi.org/10.5493/wjem.v12.i5.100

TO THE EDITOR

We read with interest a retrospective study that assessed whether serological rapid antibody tests would be effective in the diagnosis of coronavirus disease 2019 (COVID-19) pneumonia in patients whose reverse transcription-polymerase chain reaction (RT-PCR) tests were negative, despite having radiological and clinical features consistent with this condition[1]. The authors evaluated and reported the clinical aspects, laboratory results, and radiological findings of 80 suspected COVID-19 patients who had at least two negative consecutive RT-PCR tests and underwent rapid serological antibody testing. In this sense, Colloidal Gold severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) IgG/IgM Rapid Test (Beijing Hotgen Biotech Co., Ltd) was used, which is a lateral flow chromatographic immunoassay detecting total antibodies produced against the SARS-CoV-2. Therefore, the specific serological total IgM/IgG antibodies against SARS-CoV-2 were detected in 22 of these patients. The authors, then, concluded that rapid serological antibody tests may be a suitable alternative in the diagnosis of suspected COVID-19 cases, especially in highly suspected cases with negative RT-PCR results.

Regarding COVID-19 diagnosis, nucleic acid amplification tests are considered as the most sensitive ones, with RT-PCR being the gold standard method, with an overall sensitivity of 0.96 (95% confidence interval [CI]: 0.93-0.98) and false negative rate of 0.06 (95%CI: 0.04-0.08), according to a recent metaanalysis[2]. On the other hand, chest CT scan is another fundamental piece for the diagnosis of COVID-19 and monitoring of the evolution of the patient's condition[3]. Although the identification of typical lesions caused by SARS-CoV-2 is relevant, presenting a high sensitivity, it has a low specificity, since imaging findings may also be present in other viral infections with similar ongoing symptoms to COVID-19[4].

In this sense, serological tests emerged in the SARS-CoV-2 pandemic to diagnose the infection after 14 d, since this is the cut-off period for reliable detection of amplification methods [5]. One study analyzed samples of SARS-CoV-2-positive patients by RT-PCR test, SARS-CoV-2 RT-PCR-negative patients with a clinical picture of COVID-19, and controls. General sensitivity for IgG was around 80.0% for the chemiluminescence enzyme immunoassays (CLIA), enzyme-linked immunosorbent assays (ELISA), and lateral flow immunoassays (LFIA) and the sensitivity of IgG reached 100.0% when the blood was obtained 15 d after the symptoms appeared. Overall, IgG specificity was ≥ 95.8%. In addition, the same study identified an IgM sensitivity of 81.8% and specificity of 95.3% in LFIA, which were 100% after 15 d of symptom onset[6]. Otherwise, in a meta-analysis study, the authors verified the pooled sensitivity and specificity of IgG and IgM of the above cited tests and observed wide 95%CIs, varying from 46.2% to 100% (CLIA), 75.6% to 90.9% (ELISA), and 49.3% to 79.3% (LFIA), which led the authors to emphasize that the data do not support the continued use of existing point-of-care serological tests and that further studies are needed to assess the accuracy of serological tests[7].

Another meta-analysis study by analyzing RT-PCR, immunological tests, and computed tomography (CT) demonstrated that the combination of IgM and IgG antibodies yielded a sensitivity of 84.5% and specificity of 91.6%, the RT-PCR test in sputum samples and CT obtained a sensitivity of 97.2% and 91.9%, respectively, but CT had a low specificity (25.1%). The authors corroborated the consensus of the RT-PCR method being the gold standard, but recommended the combination of different tests to improve the sensitivity and specificity of the diagnosis[8].

In respect to our study, the experience with EDI™ Novel Coronavirus COVID-19 ELISA Kit Flyer IgM and IgG (Epitope diagnosis Inc São Diego, EUA) qualitative test differs from the conclusion of Yıldırım et al[1]. Our team compared a quantitative ELISA test that detects antibodies against the SARS-CoV-2 S1 epitope with the EDI™ Novel Coronavirus COVID-19 ELISA Kit Flyer IgM and IgG (Epitope Diagnosis Inc San Diego, USA), which is a qualitative test, that is, it indicates the presence or absence of the virus without quantifying the viral load[9]. Eighty Brazilian patients were included in this study (47 adults, mean age of 41.5 ± 12.2 , and 33 children, mean age of 9.7 ± 2.9), and among them, 21 were RT- PCR positive for COVID-19 and 59 were negative.

Overall, our results demonstrated that the sensitivity, specificity, accuracy, positive predictive values and negative predictive values of IgM detection were 19.05%, 100.0%, 78.7%, 100.0% and 77.6%, respectively, whereas the corresponding values of IgG were 38.1%, 100.0%, 83.7%, 100.0% and 81.9%, respectively. Notably, four children included in our study had severe multisystem inflammatory syndrome (MIS-C), which in most cases is a post-acute manifestation of COVID-19. Among the four children with MIS-C, two were RT-PCR negative, IgM was not detected in the serum of these children, but IgG was positive in three of them. Therefore, more accurate tests are necessary, not only to improve the diagnosis of COVID-19, but also of MIS-C especially because the direct detection of SARS-CoV-2 is less frequent in this severe disease. It is worth mentioning that, as shown in other studies, when comparing a quantitative ELISA test with a qualitative test, the sensitivity was much higher in the first one, even without differences in the duration of time from the onset of the first symptoms and blood collection (data not shown).

To conclude, despite the putative benefit of qualitative antibody tests in diagnosing COVID-19 in patients in whom RT-PCR test was negative, the low sensibility of some testing kits limits their use as a first-line diagnostic tool. Thus, we suggest qualitative tests to be used as an adjunctive tool in specific situations, of note: (1) In patients whose clinical picture indicates COVID-19, yet RT-PCR is negative; and (2) In the identification of past infections, until advances in the field improve the performance of rapid tests or further studies clarify the divergent results regarding the sensibility and specificity of these diagnostic methods.

FOOTNOTES

Author contributions: All authors equally contributed to this paper with conception and design of the study, literature review and analysis, manuscript drafting, critical revision, and editing, and approval of the final version.

Conflict-of-interest statement: All authors declare that they have no conflict of interest.

Open-Access: This article is an open-access article that was selected by an in-house editor and fully peer-reviewed by external reviewers. It is distributed in accordance with the Creative Commons Attribution NonCommercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited and the use is noncommercial. See: https://creativecommons.org/Licenses/by-nc/4.0/

Country/Territory of origin: Brazil

ORCID number: Fabrício Freire de Melo 0000-0002-5680-2753; Lilian Martins Oliveira Diniz 0000-0003-1410-003X; José Nélio Januário 0000-0002-0912-3782; Rafael Santos Dantas Miranda Dórea 0000-0002-5347-6913; Breno Bittencourt de Brito 0000-0002-1831-7909; Hanna Santos Marques 0000-0001-5741-1570; Fabian Fellipe Bueno Lemos 0000-0002-4686-7086; Marcel Silva Luz 0000-0003-1650-5807; Samuel Luca Rocha Pinheiro 0000-0002-8877-892X; Dulciene Maria de Magalhães Queiroz 0000-0003-1334-9423.

S-Editor: Wang LL L-Editor: Wang TQ P-Editor: Wang LL

REFERENCES

- Yıldırım F, Gulhan PY, Diken ÖE, Capraz A, Simsek M, Yildirim BB, Tavsi MR, Ozturk SY, Demirtas N, Ergil J, Dirican A, Uzar T, Karaman I, Ozkaya S. Role of serological rapid antibody test in the management of possible COVID-19 cases. World J Exp Med 2021; 11: 44-54 [PMID: 34616666 DOI: 10.5493/wjem.v11.i4.44]
- 2 Pu R, Liu S, Ren X, Shi D, Ba Y, Huo Y, Zhang W, Ma L, Liu Y, Yang Y, Cheng N. The screening value of RT-LAMP and RT-PCR in the diagnosis of COVID-19: systematic review and meta-analysis. J Virol Methods 2022; 300: 114392 [PMID: 34856308 DOI: 10.1016/j.jviromet.2021.114392]
- Ai T, Yang Z, Hou H, Zhan C, Chen C, Lv W, Tao Q, Sun Z, Xia L. Correlation of Chest CT and RT-PCR Testing for Coronavirus Disease 2019 (COVID-19) in China: A Report of 1014 Cases. Radiology 2020; 296: E32-E40 [PMID: 32101510 DOI: 10.1148/radiol.2020200642]
- Mair MD, Hussain M, Siddiqui S, Das S, Baker A, Conboy P, Valsamakis T, Uddin J, Rea P. A systematic review and meta-analysis comparing the diagnostic accuracy of initial RT-PCR and CT scan in suspected COVID-19 patients. Br J Radiol 2021; 94: 20201039 [PMID: 33353381 DOI: 10.1259/bjr.20201039]
- 5 Tang YW, Schmitz JE, Persing DH, Stratton CW. Laboratory Diagnosis of COVID-19: Current Issues and Challenges. J Clin Microbiol 2020; 58 [PMID: 32245835 DOI: 10.1128/JCM.00512-20]
- Nicol T, Lefeuvre C, Serri O, Pivert A, Joubaud F, Dubée V, Kouatchet A, Ducancelle A, Lunel-Fabiani F, Le Guillou-Guillemette H. Assessment of SARS-CoV-2 serological tests for the diagnosis of COVID-19 through the evaluation of three

102



- immunoassays: Two automated immunoassays (Euroimmun and Abbott) and one rapid lateral flow immunoassay (NG Biotech). J Clin Virol 2020; 129: 104511 [PMID: 32593133 DOI: 10.1016/j.jcv.2020.104511]
- 7 Lisboa Bastos M, Tavaziva G, Abidi SK, Campbell JR, Haraoui LP, Johnston JC, Lan Z, Law S, MacLean E, Trajman A, Menzies D, Benedetti A, Ahmad Khan F. Diagnostic accuracy of serological tests for covid-19: systematic review and metaanalysis. BMJ 2020; **370**: m2516 [PMID: 32611558 DOI: 10.1136/bmj.m2516]
- Böger B, Fachi MM, Vilhena RO, Cobre AF, Tonin FS, Pontarolo R. Systematic review with meta-analysis of the accuracy of diagnostic tests for COVID-19. Am J Infect Control 2021; 49: 21-29 [PMID: 32659413 DOI: 10.1016/j.ajic.2020.07.011]
- Ravi N, Cortade DL, Ng E, Wang SX. Diagnostics for SARS-CoV-2 detection: A comprehensive review of the FDA-EUA COVID-19 testing landscape. Biosens Bioelectron 2020; 165: 112454 [PMID: 32729549 DOI: 10.1016/j.bios.2020.112454]





Published by Baishideng Publishing Group Inc

7041 Koll Center Parkway, Suite 160, Pleasanton, CA 94566, USA

Telephone: +1-925-3991568

E-mail: bpgoffice@wjgnet.com

Help Desk: https://www.f6publishing.com/helpdesk

https://www.wjgnet.com

