



Experience in clinical application of electrogastrogram

Jian-Sheng Li

Jian-Sheng Li, Digestive Department Taiyuan Central Hospital 1 Eastern 3rd,
Taiyuan, Shanxi Province, China

Author contributions: The author solely contributed to the work.

Original title: *China National Journal of New Gastroenterology* (1995-1997)
renamed *World Journal of Gastroenterology* (1998-).

Received: December 11, 1995

Revised: February 26, 1996

Accepted: August 1, 1996

Published online: September 15, 1996

Abstract

AIM: It has been nearly 10 years to diagnose and treat gastro-pathy with the aid of electrogastrogram. Comparing with gastro-photography and gastrofiberscope in thirty thousand cases, electro-gastrogram is one of the important methods on research for disorder

of gastric motility. It also supplies certain information for gastric organic diseases. In order to improve the electrogastrography, 1 put forward proposals below: (1) Research and production of the apparatus should be integrated with clinical research. (2) According to experiences of clinical application, it is the best choice to produce single conductor electrogastrograph instead of intelligent type. (3) Only one explore electrode is needed. Routine test is only used before meal. For the observing norms. We should put stress on gastric rhythm. And (4) Electrogastrography is mostly applied to diagnosis and treatment of disorder of gastric motility.

Key words: Clinical application; Electrogastrogram

© **The Author(s) 1996.** Published by Baishideng Publishing Group Inc. All rights reserved.

Li JS. Experience in clinical application of electrogastrogram. *World J Gastroenterol* 1996; 2(Suppl1): 100 Available from: URL: <http://www.wjgnet.com/1007-9327/full/v2/iSuppl1/100.htm> DOI: <http://dx.doi.org/10.3748/wjg.v2.iSuppl1.100>

E- Editor: Liu WX



Published by **Baishideng Publishing Group Inc**
8226 Regency Drive, Pleasanton, CA 94588, USA
Telephone: +1-925-223-8242
Fax: +1-925-223-8243
E-mail: bpgoffice@wjgnet.com
Help Desk: <http://www.wjgnet.com/esps/helpdesk.aspx>
<http://www.wjgnet.com>

