

Coronary computed tomography angiography in coronary artery disease

Zhonghua Sun, Kwan-Hoong Ng

Zhonghua Sun, Discipline of Medical Imaging, Department of Imaging and Applied Physics, Curtin University, GPO Box U1987, Perth, Western Australia 6845, Australia

Kwan-Hoong Ng, Department of Biomedical Imaging, University of Malaya, Kuala Lumpur 50603, Malaysia

Kwan-Hoong Ng, University of Malaya Research Imaging Centre, University of Malaya, Kuala Lumpur 50603, Malaysia

Author contributions: All authors contributed equally to this paper.

Correspondence to: Zhonghua Sun, Associate Professor, Discipline of Medical Imaging, Department of Imaging and Applied Physics, Curtin University, GPO Box U1987, Perth, Western Australia 6845, Australia. z.sun@curtin.edu.au

Telephone: +61-8-92667509 Fax: +61-8-92662377

Received: April 29, 2011 Revised: July 6, 2011

Accepted: July 13, 2011

Published online: September 26, 2011

Abstract

AIM: To investigate the research directions of coronary computed tomography (CT) angiography in the diagnosis of coronary artery disease (CAD) based on a systematic review of the literature.

METHODS: A search of articles on coronary CT angiography in the diagnosis of CAD was performed during a 6-year-period between 2005 and 2010 from five main radiology journals namely, Radiology, American Journal of Roentgenology, European Radiology, European Journal of Radiology and British Journal of Radiology. Analysis of the references was focused on the research directions of coronary CT angiography with regard to the type of studies in terms of diagnostic value, application of dose-reduction strategies and resultant effective radiation doses with use of these techniques.

RESULTS: One hundred and forty two studies were identified which met the selection criteria and were included in the analysis. 64-slice CT (single source and

dual-source CT) dominated 78% of the coronary CT angiography studies. Prior to 2007, research was focused on the diagnostic value of coronary CT angiography, but since 2008 more attention has been paid to radiation dose reduction. Radiation dose was reported in 64 studies, representing 45% of total studies published in the five radiology journals. Various dose-saving strategies have been implemented and prospective electrocardiography-triggering and high pitch techniques were found to be the most effective approaches for radiation dose reduction, with the corresponding mean effective dose being 3.5 ± 1.9 mSv and 1.7 ± 0.6 mSv, respectively.

CONCLUSION: This review shows that the current research in coronary CT angiography has shifted from the previous focus on diagnostic accuracy in CAD to more emphasis on radiation dose reduction.

© 2011 Baishideng. All rights reserved.

Key words: Coronary artery disease; Diagnostic value; Multislice computed tomography; Radiation dose; Risk

Peer reviewers: Ole Dyg Pedersen, MD, Department of Cardiology, Bispebjerg University Hospital, 2400 Copenhagen, Denmark; Enrique Vallejo, MD, FASN, Instituto Nacional de Cardiologia Ignacio Chavez, Juan Badiano No.1 Col. Seccion XVI, Tlalpan 14080, Mexico DF, Mexico

Sun Z, Ng KH. Coronary computed tomography angiography in coronary artery disease. *World J Cardiol* 2011; 3(9): 303-310 Available from: URL: <http://www.wjgnet.com/1949-8462/full/v3/i9/303.htm> DOI: <http://dx.doi.org/10.4330/wjc.v3.i9.303>

INTRODUCTION

Coronary artery disease (CAD) is the leading cause of morbidity and mortality in many advanced countries and its prevalence is increasing in developing countries^[1,2].

According to recent World Health Organization statistics for 2007, cardiovascular deaths account for 33.7% of all deaths worldwide, whereas cancer represents 29.5%, other chronic diseases 26.5%, injury 7%, and communicable diseases 4.6%^[3]. CAD is the leading cause of cardiovascular death throughout the world. In light of the current global focus on healthcare utilization, costs, and quality, it is essential to monitor and understand the magnitude of healthcare delivery and costs, as well as the quality of healthcare delivery in relation to the CAD.

Invasive coronary angiography (ICA) is widely used as a reliable technique to diagnose CAD because of its superior spatial and temporal resolution. However, it is an invasive and expensive procedure with an associated small percentage of morbidity and mortality^[4]. As an alternative to ICA, coronary computed tomography (CT) angiography has been increasingly used for the investigation of suspected CAD, and rapid technological developments of multislice CT (MSCT) have led to both improved spatial and temporal resolution^[5-7]. Studies have shown that coronary CT angiography has a high diagnostic accuracy for the detection of significant CAD ($\geq 50\%$ lumen stenosis) when compared to ICA^[8-11], and in selected patients, coronary CT angiography is recommended as a reliable alternative to ICA in the diagnosis of CAD.

Despite promising results having been achieved with coronary CT angiography, it has the disadvantage of high radiation dose, which leads to the concern of radiation-associated risks^[12,13]. It is generally agreed that CT is an imaging modality with high radiation exposure, as it contributes up to 70% of the radiation dose of all radiological examinations, although it comprises only 15% of all radiological examinations. Radiation-induced malignancy is a problem that has raised serious concern in the medical field and this has been addressed by the National Research Council of the United States^[14]. It is reported that radiation dose from a CT scan has been significantly underestimated by both radiologists and physicians^[12,15]. Despite the increased awareness of radiation risk, many clinicians and researchers have not realized the amount of radiation exposure associated with coronary CT angiography, or the possibility of tailoring the scanning protocols to reduce radiation exposure to patients. Thus, the purpose of this article was to investigate the current research directions of coronary CT angiography in CAD, based on a systematic review of the literature published in five main radiology journals over the past 6 years.

MATERIALS AND METHODS

References searching

A search of articles on coronary CT angiography in cardiac imaging was performed during a 6-year-period between 2005 and 2010 in five main radiology journals, *Radiology*, *American Journal of Roentgenology (AJR)*, *European Radiology (ER)*, *European Journal of Radiology (EJR)* and *British Journal of Radiology (BJR)*. The reason we started

our search in 2005 is because 64-slice CT was introduced in 2004. The keywords used in searching the references in each journal website included: MSCT in CAD, diagnostic value of MSCT angiography in CAD, coronary CT angiography and CT coronary angiography, prospective electrocardiography (ECG)-gating/triggering in cardiac imaging, and cardiac CT with radiation dose reduction. In addition, a search for relevant articles was conducted by electronically checking the monthly issue of each journal's publications, and eligible full-text articles were obtained for data extraction and analysis. Prospective and retrospective studies were included if they met the following criteria: (1) **studies conducted on human subjects** or phantom with use of MSCT or coronary CT angiography in CAD; (2) **either coronary calcium scoring, diagnosis of CAD, prognostic value of CAD, or myocardial perfusion imaging** must be addressed in each study; (3) coronary CT angiography in terms of prospective ECG-triggering or retrospective gating must be clearly stated in each study; and (4) **any strategies to reduce radiation dose** must be provided. Since it is possible that many studies would not meet the third and fourth criteria, studies were still eligible for inclusion in the analysis as long as they met the first two criteria. Exclusion criteria were: studies on coronary stenting or coronary artery-bypass surgery; review articles or case reports; a letter or comment to the editor.

Data extraction and analysis

All articles were reviewed and extracted by two independent authors with regard to study design and procedure techniques. Each author independently assessed the retrieved articles according to the selection criteria and inter-observer agreement was reached by consensus. The authors looked for the following characteristics in each study: year of publication; type of scanning unit used for coronary CT angiography; type of studies in terms of prospective or retrospective studies; researcher background, e.g. radiologists, cardiologists, physicians or other health care professionals (for example, medical imaging specialists); number of research centres where studies were performed. In addition, more focus was placed on the research directions of coronary CT angiography with the aim of seeking information on the application of dose-saving strategies. This included the common approaches for dose reduction: adjustments of kVp and mAs (tube current modulation, either anatomy-based or ECG-controlled tube current modulation); high pitch value; prospective ECG-triggering *vs* retrospective ECG-gating.

Statistical analysis

All variables are expressed as mean \pm SD. Statistical tests were performed using SPSS V 18.0 (SPSS, Inc., Chicago, IL, United States). Comparisons were performed using one sample *T* test. A *P* value less than 0.05 was defined as statistically significant difference.

RESULTS

General information

One hundred and forty-two studies were identified which met the selection criteria and were included in the analysis. One hundred and thirty seven studies were conducted at a single centre, while three studies were performed at 3 centres and the remaining two studies at 9 and 20 study sites, respectively. The majority of these studies were published in *ER* with the highest number being 62 articles, and the least number of studies was found in the *BJR* with only 11 studies being identified. For the remaining three journals, *Radiology*, *AJR* and *EJR*, the corresponding studies were 23, 24 and 22, respectively.

Analysis of author background indicated that most of the studies were performed by radiologists, followed by cardiologists, and other health care professionals such as physicians and medical imaging specialists. Radiologists accounted for 84% of the authors who performed the studies, while cardiologists were in collaboration with radiologists as a research team in all of the studies, except in two studies published in *ER*, where cardiologists alone conducted the research.

Of 142 studies, 126 were performed on patients (89%), and the remaining 16 studies were conducted on anthropomorphic phantoms with the aim of investigating the effect of heart rate on diagnostic image quality, or radiation dose or high pitch value on image quality. 71 studies (50%) were performed to study the diagnostic value (67 studies) and prognostic value (4 studies) of coronary CT angiography in CAD, while the remaining studies were mainly focused on radiation dose reduction.

Type of MSCT scanners

Variable MSCT scanners were used in these studies, with 64- or more slice-scanners representing the majority (85%) of the studies. Figure 1 shows the distribution of different generations of MSCT scanners that were identified in the analysis. Despite rapid technological developments of MSCT scanners, such as the increased availability of 256- and 320-slice CT, single source 64-slice and dual-source CT (DSCT) still dominated 78% of the coronary CT angiography studies. Coronary CT angiography was compared with integrated single photon emission computed tomography/CT and positron emission tomography/CT in two studies to investigate the myocardial perfusion value of coronary CT angiography.

Research directions of coronary CT angiography

The issue of radiation dose has been addressed since 2007 in the publications, *Radiology*, *AJR* and *BJR*, while this has been addressed since 2008 in the *EJR* publications. In comparison, radiation dose was first reported in some studies published in 2006 in the journal of *ER*, although more attention has been paid to the issue of dose since 2007. Prior to 2006 and 2007, research was focused on the diagnostic value of coronary CT angiography in CAD, regardless of the studies performed with 8-slice,




Figure 1 The number of studies performed with different generations of multislice computed tomography scanners. CT: Computed tomography; DSCT: Dual-source CT; MSCT: Multislice CT; PET: Positron emission tomography; SPECT: Single photon emission computed tomography.

16- or 64-slice CT scanners. From 2008, more and more studies started to address radiation dose reduction in these five radiology journals.

Dose-saving strategies

Various dose-saving strategies have been implemented in these studies, and have included adjustments of tube voltage (kVp), tube current (tube current modulation), pitch value, prospective ECG-triggering *vs* retrospective ECG-gating, addition of padding in patients with high heart rate variability, and application of adaptive statistical iterative reconstruction (ASIR) algorithms for image noise reduction. Radiation dose was reported in 64 studies, representing 45% of the total studies published in the five radiology journals. Of these 64 studies, nearly 50% were published in the journal of *ER*.

Prospective ECG-triggering was the most commonly applied technique for dose reduction, with 25 studies reported in these five radiology journals. Of these 25 studies, a direct comparison between prospective ECG-triggering and retrospective gating was performed in 16 studies. The mean effective dose was 3.5 ± 1.9 mSv for prospective ECG-triggering, which is significantly lower ($P < 0.0001$) than that acquired with retrospective ECG-gating, which is 16.3 ± 6.3 mSv.

ECG-controlled tube current modulation was the second most common approach utilized in 22 studies for dose reduction. However, the effective dose value was only reported in 7 studies, with the mean dose being 8.8 ± 2.6 mSv. Application of pitch (up to 3.4) in coronary CT angiography was reported in five studies with the mean effective dose being 1.7 ± 0.6 mSv. Lower kVp of 100 *vs* the standard 120 kVp was compared in four studies with the resultant mean effective dose of 5.4 ± 2.9 mSv for 100 kVp protocols, and 9.4 ± 5.1 mSv for 120 kVp protocols, leading to a dose reduction of 42%. Figure 2 shows the mean effective dose associated with the above-mentioned commonly used strategies for dose reduction. As shown in the figure, prospective ECG-trig-




Figure 2 Box plot shows the mean effective dose associated with different dose-reduction techniques. Coronary computed tomography angiography with the use of high pitch and prospective electrocardiography-triggering leads to the lowest radiation dose. The boxes indicate the first to third quartiles, and each midline indicates the median (second quartile) and the whiskers represent the maximum and minimum values of effective dose.

gating and high pitch resulted in the lowest doses when compared to other dose-saving strategies.

ASIR was reported in two studies and the mean effective dose was less than 3.0 mSv. A comparison of radiation dose with use of the International Commission on Radiological Protection (ICRP) 103 and ICRP 60 documents was conducted in two studies with a significant difference observed, indicating that the radiation dose was underestimated if dose calculation is based on the weighting factors that were reported in the ICRP 60.

Radiation dose corresponding to different heart rates was reported in three studies and the analysis shows that DSCT is superior to single-source CT in the diagnosis of CAD in patients with higher heart rates, with resultant lower doses. Padding was applied in one study performed with prospective ECG-gating. The purpose of adding padding is to provide additional phase information to compensate for variations in heart rate by adding time before and after the centre phase of the acquisition. Padding is described in the range of 0-200 ms and is added to both sides of the centre of the acquisition with padding 0 corresponding to a window of 100 ms scanning time and padding 100 corresponding to a window of 200 ms scanning time. In that study, the effective dose was 2.3 mSv for 0 padding, and the dose increased to 3.8 mSv for 1-99 padding and to 5.5 mSv for 100-150 padding.

The remaining study involved the investigation of the effect of adjusting the scan length of coronary CT angiography using the calcium scoring images instead of the scout view with regard to radiation dose. The effective dose associated with the calcium scoring-derived length (9.0 ± 0.6 mSv) was significantly lower than that using the scout view-derived length (10.7 ± 1.2 mSv), corresponding to a radiation dose reduction of 16%.

DISCUSSION

This analysis presents three findings which are considered

important for the clinical application of coronary CT angiography in CAD. Firstly, there is a change in research direction with a shift of focus from the previous diagnostic value of coronary CT angiography in CAD to an emphasis on radiation dose reduction. This indicates the increased awareness of radiation dose to patients during coronary CT angiography examinations. Secondly, more dose-saving strategies have been implemented to reduce radiation dose associated with coronary CT angiography while still maintaining diagnostic image quality. Of these strategies, prospective ECG-triggering and high pitch protocols represent the most effective approaches with a significant reduction of radiation dose compared to conventional retrospective gating or other approaches. Thirdly, coronary CT angiography performed with 64-slice CT (single-source or DSCT) dominated the majority of studies that were reviewed, and this indicates that 64-slice CT still plays a key role in the clinical diagnosis of CAD, despite the emergence of 256- and 320-slice CT.

Over the last decade a great deal of interest has been focused on the imaging and diagnosis of CAD using coronary CT angiography, due to its less invasive nature and improved spatial and temporal resolution. Moderate to high diagnostic accuracy was achieved with 64- or more slice CT, owing to further technical improvements^[16-20]. These studies have indicated that coronary CT angiography has high accuracy in the diagnosis of CAD and could be used as an effective alternative to ICA in selected patients. However, coronary CT angiography has the disadvantage of high radiation dose which raises concerns for both clinicians and manufacturers. This is reflected in the changing research directions from the early research focus on the diagnostic value of coronary CT angiography in CAD to the increasingly reported studies on dose reduction, as demonstrated in this analysis.

Radiation dose is becoming a major issue for coronary CT angiography, since 64- or more-slice CT shows improved and promising results in the diagnosis of CAD^[21-25]. The radiation risks associated with coronary CT angiography have raised serious concerns and have become a hot topic of debate in the literature^[12,13,26]. The general view about radiation dose is that CT is associated with a risk of cancer development. Therefore, coronary CT angiography should be performed with dose-saving strategies whenever possible to reduce the radiation dose to patients.

One of the most effective approaches for dose reduction is adjustment of the tube current according to ECG signal, which is defined as ECG-controlled tube current modulation. This approach represents the most significant improvement in minimizing radiation exposure from coronary CT angiography. It has been reported that radiation dose can be reduced by 30%-50% through modulation of the tube current output to decrease the dose given during the systolic phase^[27,28]. The estimated radiation dose reduction is similar to or less than that of an ICA examination with use of this dose saving strategy^[27,29]. The mean effective dose of ECG-controlled tube current modulation in

this analysis was 8.8 mSv, and this was comparable to that of ICA, which was between 3-9 mSv^[30].

Prospective ECG-triggering has been confirmed to be another efficient technique for radiation dose reduction in coronary CT angiography^[31-33]. Use of prospective ECG-triggering with 64-slice or DSCT has been reported to reduce the effective radiation dose by up to 90% when compared to retrospective ECG-gating, with diagnostic image quality being achieved in more than 90% of cases^[34-39]. This was confirmed in this analysis, as the mean effective dose of prospective triggering was 3.5 mSv, which is equivalent to or even lower than that of ICA.

Another effective method currently undertaken in coronary CT angiography to reduce radiation dose is to lower the tube voltage, since radiation dose varies with the square of the kV. Modern CT scanners include tube voltages of 120 or 140 kV, reflecting the settings which most often result in adequate image quality. However, acquisition of cardiac CT images with 100 kV, or even lower, is possible and has been suggested as an effective means to reduce radiation dose in coronary CT angiography^[40,41]. Up to 70% dose reduction has been reported in the literature when the X-ray tube voltage is decreased from 120 kVp to 100 kVp or even 80 kVp^[42,43]. This analysis showed that a 42% dose reduction was observed in coronary CT angiography with the use of 100 kVp when compared to the routine 120 kVp protocol, and this is consistent with the results reported in the literature. Therefore, tube voltage can be adjusted in cardiac CT angiography without affecting diagnostic image quality, and this should be applied whenever possible in clinical practice^[44,45].

For coronary CT angiography, very low pitch values (0.2-0.4) are typically required for coronary data acquisition to ensure continuous z-axis coverage between image stacks reconstructed from consecutive cardiac cycles. The main disadvantage of this approach is higher radiation exposure since the table is advanced by less than one detector width during each gantry rotation, thus, the same anatomic area is exposed to X-ray radiation during consecutive rotations of the gantry. Increasing the pitch to a higher value was made possible by the development of the second generation of DSCT scanners, Siemens Definition Flash, which enabled acquisition of 128 slices simultaneously (flying focal spot)^[46-50]. This DSCT mode allows coronary CT angiography to be performed at high pitch values of up to 3.4 with a significant reduction in radiation dose. A high pitch was applied in five studies in this analysis with a mean effective dose of 1.72 mSv, and a radiation dose less than 1.0 mSv was also reported in one study, indicating the effectiveness of this dose-reduction approach. The new scan mode, with a superior temporal resolution of 75 ms, is regarded as an attractive alternative to ICA due to the very low dose and high image quality, although more studies based on a large cohort are required.

Although nearly 50% of the studies in this analysis were conducted with the use of single-source CT, DSCT was used in about one-third of the studies, indicating the increased use of this improved technique. The main

limitation of single-source CT coronary angiography is that β -blockers are frequently used to lower the heart rate to less than 65 beats per minute. The introduction of DSCT marked another technological improvement of MSCT in cardiac imaging, as the temporal resolution was further increased from 165 ms to 83 ms, thus eliminating the need to control the heart rate during the scan^[43,44,51]. Studies comparing DSCT with single-source CT demonstrated that DSCT maintains high diagnostic accuracy in the diagnostic examination of a wide range of patient subsets, e.g., patients with higher and even irregular heart rates (including atrial fibrillation)^[52-54]. Despite slightly lower per-segment evaluability in patients with higher heart rates, DSCT did not show decreased diagnostic accuracy for the detection of coronary stenoses^[54,55]. DSCT improves temporal resolution which is vital in patients who cannot have β blockade.

Expansion of MSCT systems from a 64-slice to 256- and 320-slice system has allowed for the accurate assessment of stenosis severity and atherosclerotic plaque composition, or even the acquisition of whole-heart coverage in one gantry rotation, thus, eliminating the restrictions and limitations associated with 64-slice CT scanners^[22,24]. Studies performed with 256- and 320-slice CT demonstrated the high diagnostic value of coronary CT angiography for the quantification of coronary lumen stenosis and assessment of plaque morphology and distribution in the coronary arteries^[22,56,57]. In patients with atrial fibrillation, 320-slice CT allows for excellent visualization of all coronary segments with sufficient images to enable a diagnosis, and acquisition of images with consistent CT attenuation along the different coronary arteries^[23,57]. Despite these advantages, 256- and 320-slice CT has the limitations of reduced gantry speed (270-350 ms), high radiation dose and cone beam effects which need to be addressed in future technical developments.

Some limitations exist in this review. Firstly, our search for references only focused on the five main radiology journals. However, research articles on cardiac CT imaging are also published in cardiac journals such as *European Heart Journal*, *Journal of American College of Cardiology*, *Circulation*, *American Journal of Cardiology* and *International Journal of Cardiovascular Imaging*. Therefore, it is possible that some relevant references were not included in this analysis. Secondly, although tube current modulation is one of the most effective methods for radiation dose reduction, a number of studies did not report the actual effective dose, despite the application of this technique in their studies. This emphasises the importance of awareness of radiation exposure to patients by physicians; thus, radiation dose values (effective dose or dose length product) should be recorded in each study to enable the comparison of dose-saving techniques between different studies. Thirdly, we tried to include as many references as possible that were available over the last few years; however, some articles which were accepted for publication were excluded from the analysis. Finally, although this analysis shows that 64-slice CT coronary

angiography dominated the majority of the studies, there is no doubt that more and more studies are being performed with the latest models such as 256- and 320-slice CT scanners. Further research to include these studies is necessary.

In conclusion, this systematic review of the literature from five main radiology journals shows that the current research in coronary CT angiography has shifted from the previous focus on diagnostic accuracy in CAD to more emphasis on radiation dose reduction. This review also indicates the increased awareness of radiation dose associated with coronary CT angiography in the literature. Various dose-saving strategies have been undertaken in the past few years to lower radiation exposure to patients undergoing coronary CT angiography. Effective dose reduction has been achieved by employing techniques with a radiation dose of around 10 mSv to as low as 1 mSv in some studies. It is important to note that the current effective doses from coronary CT angiography are at the same level or even lower than those acquired from ICA. Therefore, according to this analysis, the achievements in radiation dose reduction in coronary CT angiography have been tremendous.

COMMENTS

Background

Coronary computed tomography (CT) angiography is increasingly used to diagnose coronary artery disease (CAD) due to its reduced invasiveness and improved diagnostic accuracy. Despite the high diagnostic value of coronary CT angiography, it has the disadvantage of high radiation dose which contributes to radiation-induced malignancy. This study was conducted to investigate the current research directions of coronary CT angiography in the diagnosis of CAD based on a systematic review of the radiology literature.

Research frontiers

Current research in coronary CT angiography has shifted from the previous focus on diagnostic accuracy in CAD to more emphasis on radiation dose reduction. This reflects the increased awareness of radiation dose associated with coronary CT angiography in the literature.

Innovations and breakthroughs

Various dose-saving strategies have been proposed in the literature with regard to radiation dose reduction associated with coronary CT angiography. This study shows that these dose-saving strategies have been very effective in dose reduction, and it is possible to lower the level of radiation dose to that of invasive coronary angiography without compromising diagnostic image quality.

Applications

Coronary CT angiography is mainly performed with the use of 64-slice CT, however, with technological developments in CT scanning techniques, the latest models such as 256- and 320-slice CT are increasingly used to accurately diagnose patients with suspected CAD due to whole-heart coverage in one gantry rotation.

Terminology

Effective dose is a single parameter used to reflect the relative risk of exposure to ionizing radiation. Calculation of the effective dose of coronary CT angiography takes into account the biological effect of the radiation on the heart, because each organ is given a tissue weighting depending on its individual susceptibility to the effects of ionizing radiation.

Peer review

This paper reviews recent developments in technology of CT-angiography of the coronary arteries and elude on future directions in development of this technology. The paper covering this relatively new non-invasive technology is in general interest for the readers of the journal.

REFERENCES

- 1 **Lloyd-Jones D**, Adams RJ, Brown TM, Carnethon M, Dai S, De Simone G, Ferguson TB, Ford E, Furie K, Gillespie C, Go A, Greenlund K, Haase N, Hailpern S, Ho PM, Howard V, Kissela B, Kittner S, Lackland D, Lisabeth L, Marelli A, McDermott MM, Meigs J, Mozaffarian D, Mussolino M, Nichol G, Roger VL, Rosamond W, Sacco R, Sorlie P, Stafford R, Thom T, Wasserthiel-Smoller S, Wong ND, Wylie-Rosett J. Executive summary: heart disease and stroke statistics--2010 update: a report from the American Heart Association. *Circulation* 2010; **121**: 948-954
- 2 **Gaziano TA**, Bitton A, Anand S, Abrahams-Gessel S, Murphy A. Growing epidemic of coronary heart disease in low- and middle-income countries. *Curr Probl Cardiol* 2010; **35**: 72-115
- 3 **World Health Organization**. World Health Report 2002: Reducing Risks, Promoting Healthy Life. Geneva: World Health Organization, 2002
- 4 **Noto TJ**, Johnson LW, Krone R, Weaver WF, Clark DA, Kramer JR, Vetrovec GW. Cardiac catheterization 1990: a report of the Registry of the Society for Cardiac Angiography and Interventions (SCA& amp; I). *Cathet Cardiovasc Diagn* 1991; **24**: 75-83
- 5 **Nieman K**, Oudkerk M, Rensing BJ, van Ooijen P, Munne A, van Geuns RJ, de Feyter PJ. Coronary angiography with multi-slice computed tomography. *Lancet* 2001; **357**: 599-603
- 6 **Kuettner A**, Trabold T, Schroeder S, Feyer A, Beck T, Brueckner A, Heuschmid M, Burgstahler C, Kopp AF, Claussen CD. Noninvasive detection of coronary lesions using 16-detector multislice spiral computed tomography technology: initial clinical results. *J Am Coll Cardiol* 2004; **44**: 1230-1237
- 7 **Pugliese F**, Mollet NR, Runza G, van Mieghem C, Meijboom WB, Malagutti P, Baks T, Krestin GP, deFeyter PJ, Cademartiri F. Diagnostic accuracy of non-invasive 64-slice CT coronary angiography in patients with stable angina pectoris. *Eur Radiol* 2006; **16**: 575-582
- 8 **Sun Z**, Jiang W. Diagnostic value of multislice computed tomography angiography in coronary artery disease: a meta-analysis. *Eur J Radiol* 2006; **60**: 279-286
- 9 **Miller JM**, Rochitte CE, Dewey M, Arbab-Zadeh A, Niinuma H, Gottlieb I, Paul N, Clouse ME, Shapiro EP, Hoe J, Lardo AC, Bush DE, de Roos A, Cox C, Brinker J, Lima JA. Diagnostic performance of coronary angiography by 64-row CT. *N Engl J Med* 2008; **359**: 2324-2336
- 10 **Budoff MJ**, Dowe D, Jollis JG, Gitter M, Sutherland J, Halamert E, Scherer M, Bellinger R, Martin A, Benton R, Delago A, Min JK. Diagnostic performance of 64-multidetector row coronary computed tomographic angiography for evaluation of coronary artery stenosis in individuals without known coronary artery disease: results from the prospective multicenter ACCURACY (Assessment by Coronary Computed Tomographic Angiography of Individuals Undergoing Invasive Coronary Angiography) trial. *J Am Coll Cardiol* 2008; **52**: 1724-1732
- 11 **Meijboom WB**, Meijs MF, Schuijf JD, Cramer MJ, Mollet NR, van Mieghem CA, Nieman K, van Werkhoven JM, Pundziute G, Weustink AC, de Vos AM, Pugliese F, Rensing B, Jukema JW, Bax JJ, Prokop M, Doevendans PA, Hunink MG, Krestin GP, de Feyter PJ. Diagnostic accuracy of 64-slice computed tomography coronary angiography: a prospective, multicenter, multivendor study. *J Am Coll Cardiol* 2008; **52**: 2135-2144
- 12 **Brenner DJ**, Hall EJ. Computed tomography--an increasing source of radiation exposure. *N Engl J Med* 2007; **357**: 2277-2284
- 13 **Hausleiter J**, Meyer T, Hermann F, Hadamitzky M, Krebs M, Gerber TC, McCollough C, Martinoff S, Kastrati A, Schömig A, Achenbach S. Estimated radiation dose associated with cardiac CT angiography. *JAMA* 2009; **301**: 500-507
- 14 **Committee to Assess Health Risks from Exposure to Low**

- Levels of Ionizing Radiation.** National Research Council. Health Risks From Exposure to Low Levels of Ionizing Radiation: BEIR VII Phase 2. Washington, DC: The National Academies Press, 2006
- 15 Lee CI, Haims AH, Monico EP, Brink JA, Forman HP. Diagnostic CT scans: assessment of patient, physician, and radiologist awareness of radiation dose and possible risks. *Radiology* 2004; **231**: 393-398
 - 16 Sun Z, Lin C, Davidson R, Dong C, Liao Y. Diagnostic value of 64-slice CT angiography in coronary artery disease: a systematic review. *Eur J Radiol* 2008; **67**: 78-84
 - 17 Vanhoenacker PK, Heijenbrok-Kal MH, Van Heste R, Decramer I, Van Hoe LR, Wijns W, Hunink MG. Diagnostic performance of multidetector CT angiography for assessment of coronary artery disease: meta-analysis. *Radiology* 2007; **244**: 419-428
 - 18 Abdulla J, Abildstrom SZ, Gotzsche O, Christensen E, Kober L, Torp-Pedersen C. 64-multislice detector computed tomography coronary angiography as potential alternative to conventional coronary angiography: a systematic review and meta-analysis. *Eur Heart J* 2007; **28**: 3042-3050
 - 19 Mowatt G, Cook JA, Hillis GS, Walker S, Fraser C, Jia X, Waugh N. 64-Slice computed tomography angiography in the diagnosis and assessment of coronary artery disease: systematic review and meta-analysis. *Heart* 2008; **94**: 1386-1393
 - 20 Stein PD, Yaekoub AY, Matta F, Sostman HD. 64-slice CT for diagnosis of coronary artery disease: a systematic review. *Am J Med* 2008; **121**: 715-725
 - 21 Raff GL, Gallagher MJ, O'Neill WW, Goldstein JA. Diagnostic accuracy of noninvasive coronary angiography using 64-slice spiral computed tomography. *J Am Coll Cardiol* 2005; **46**: 552-557
 - 22 Chao SP, Law WY, Kuo CJ, Hung HF, Cheng JJ, Lo HM, Shyu KG. The diagnostic accuracy of 256-row computed tomographic angiography compared with invasive coronary angiography in patients with suspected coronary artery disease. *Eur Heart J* 2010; **31**: 1916-1923
 - 23 Rybicki FJ, Otero HJ, Steigner ML, Vorobiof G, Nallamshetty L, Mitsouras D, Ersoy H, Mather RT, Judy PF, Cai T, Coyner K, Schultz K, Whitmore AG, Di Carli MF. Initial evaluation of coronary images from 320-detector row computed tomography. *Int J Cardiovasc Imaging* 2008; **24**: 535-546
 - 24 Dewey M, Zimmermann E, Deissenrieder F, Laule M, Dübel HP, Schlattmann P, Knebel F, Rutsch W, Hamm B. Noninvasive coronary angiography by 320-row computed tomography with lower radiation exposure and maintained diagnostic accuracy: comparison of results with cardiac catheterization in a head-to-head pilot investigation. *Circulation* 2009; **120**: 867-875
 - 25 Paul JF, Abada HT. Strategies for reduction of radiation dose in cardiac multislice CT. *Eur Radiol* 2007; **17**: 2028-2037
 - 26 Raff GL, Chinnaiyan KM, Share DA, Goraya TY, Kazerooni EA, Moscucci M, Gentry RE, Abidov A. Radiation dose from cardiac computed tomography before and after implementation of radiation dose-reduction techniques. *JAMA* 2009; **301**: 2340-2348
 - 27 Jakobs TF, Becker CR, Ohnesorge B, Flohr T, Suess C, Schoepf UJ, Reiser MF. Multislice helical CT of the heart with retrospective ECG gating: reduction of radiation exposure by ECG-controlled tube current modulation. *Eur Radiol* 2002; **12**: 1081-1086
 - 28 Deetjen A, Möllmann S, Conradi G, Rolf A, Schmermund A, Hamm CW, Dill T. Use of automatic exposure control in multislice computed tomography of the coronaries: comparison of 16-slice and 64-slice scanner data with conventional coronary angiography. *Heart* 2007; **93**: 1040-1043
 - 29 Abada HT, Larchez C, Daoud B, Sigal-Cinqualbre A, Paul JF. MDCT of the coronary arteries: feasibility of low-dose CT with ECG-pulsed tube current modulation to reduce radiation dose. *AJR Am J Roentgenol* 2006; **186**: S387-S390
 - 30 Achenbach S, Giesler T, Ropers D, Ulzheimer S, Derlien H, Schulte C, Wenkel E, Moshage W, Bautz W, Daniel WG, Kalender WA, Baum U. Detection of coronary artery stenoses by contrast-enhanced, retrospectively electrocardiographically-gated, multislice spiral computed tomography. *Circulation* 2001; **103**: 2535-2538
 - 31 Hsieh J, Londt J, Vass M, Li J, Tang X, Okerlund D. Step-and-shoot data acquisition and reconstruction for cardiac x-ray computed tomography. *Med Phys* 2006; **33**: 4236-4248
 - 32 Schoenhagen P. Back to the future: coronary CT angiography using prospective ECG triggering. *Eur Heart J* 2008; **29**: 153-154
 - 33 Shuman WP, Branch KR, May JM, Mitsumori LM, Lockhart DW, Dubinsky TJ, Warren BH, Caldwell JH. Prospective versus retrospective ECG gating for 64-detector CT of the coronary arteries: comparison of image quality and patient radiation dose. *Radiology* 2008; **248**: 431-437
 - 34 Hirai N, Horiguchi J, Fujioka C, Kiguchi M, Yamamoto H, Matsuura N, Kitagawa T, Teragawa H, Kohno N, Ito K. Prospective versus retrospective ECG-gated 64-detector coronary CT angiography: assessment of image quality, stenosis, and radiation dose. *Radiology* 2008; **248**: 424-430
 - 35 Earls JP, Berman EL, Urban BA, Curry CA, Lane JL, Jennings RS, McCulloch CC, Hsieh J, Londt JH. Prospectively gated transverse coronary CT angiography versus retrospectively gated helical technique: improved image quality and reduced radiation dose. *Radiology* 2008; **246**: 742-753
 - 36 Pontone G, Andreini D, Bartorelli AL, Cortinovis S, Mushtaq S, Bertella E, Annoni A, Formenti A, Nobili E, Trabattoni D, Montorsi P, Ballerini G, Agostoni P, Pepi M. Diagnostic accuracy of coronary computed tomography angiography: a comparison between prospective and retrospective electrocardiogram triggering. *J Am Coll Cardiol* 2009; **54**: 346-355
 - 37 Klass O, Jeltsch M, Feuerlein S, Brunner H, Nagel HD, Walker MJ, Brambs HJ, Hoffmann MH. Prospectively gated axial CT coronary angiography: preliminary experiences with a novel low-dose technique. *Eur Radiol* 2009; **19**: 829-836
 - 38 Sun Z, Ng KH. Prospective versus retrospective ECG-gated multislice CT coronary angiography: A systematic review of radiation dose and diagnostic accuracy. *Eur J Radiol* 2011; Epub ahead of print
 - 39 Sun Z. Multislice CT angiography in cardiac imaging: prospective ECG-gating or retrospective ECG-gating? *Biomed Imaging Interv J* 2010; **6**: e4
 - 40 Bae KT, Seock BA, Hildebolt CF, Tao C, Zhu F, Kanematsu M, Woodard PK. Contrast enhancement in cardiovascular MDCT: effect of body weight, height, body surface area, body mass index, and obesity. *AJR Am J Roentgenol* 2008; **190**: 777-784
 - 41 Feuchtner GM, Jodocy D, Klauser A, Haberfellner B, Aglan I, Spoeck A, Hiehs S, Soegner P, Jaschke W. Radiation dose reduction by using 100-kV tube voltage in cardiac 64-slice computed tomography: a comparative study. *Eur J Radiol* 2010; **75**: e51-e56
 - 42 Pflederer T, Rudofsky L, Ropers D, Bachmann S, Marwan M, Daniel WG, Achenbach S. Image quality in a low radiation exposure protocol for retrospectively ECG-gated coronary CT angiography. *AJR Am J Roentgenol* 2009; **192**: 1045-1050
 - 43 Leschka S, Stolzmann P, Schmid FT, Scheffel H, Stinn B, Marincek B, Alkadhi H, Wildermuth S. Low kilovoltage cardiac dual-source CT: attenuation, noise, and radiation dose. *Eur Radiol* 2008; **18**: 1809-1817
 - 44 Alkadhi H, Stolzmann P, Scheffel H, Desbiolles L, Baumüller S, Plass A, Genoni M, Marincek B, Leschka S. Radiation dose of cardiac dual-source CT: the effect of tailoring the protocol to patient-specific parameters. *Eur J Radiol* 2008; **68**: 385-391
 - 45 Sun Z. Multislice CT angiography in coronary artery disease: Technical developments, radiation dose and diagnostic value. *World J Cardiol* 2010; **2**: 333-343
 - 46 Achenbach S, Marwan M, Ropers D, Schepis T, Pflederer T,

- Anders K, Kuettner A, Daniel WG, Uder M, Lell MM. Coronary computed tomography angiography with a consistent dose below 1 mSv using prospectively electrocardiogram-triggered high-pitch spiral acquisition. *Eur Heart J* 2010; **31**: 340-346
- 47 **Lell M**, Marwan M, Schepis T, Pflederer T, Anders K, Flohr T, Allmendinger T, Kalender W, Ertel D, Thierfelder C, Kuettner A, Ropers D, Daniel WG, Achenbach S. Prospectively ECG-triggered high-pitch spiral acquisition for coronary CT angiography using dual source CT: technique and initial experience. *Eur Radiol* 2009; **19**: 2576-2583
- 48 **Ertel D**, Lell MM, Harig F, Flohr T, Schmidt B, Kalender WA. Cardiac spiral dual-source CT with high pitch: a feasibility study. *Eur Radiol* 2009; **19**: 2357-2362
- 49 **Feuchtner G**, Goetti R, Plass A, Baumuegger S, Stolzmann P, Scheffel H, Wieser M, Marincek B, Alkadhi H, Leschka S. Dual-step prospective ECG-triggered 128-slice dual-source CT for evaluation of coronary arteries and cardiac function without heart rate control: a technical note. *Eur Radiol* 2010; **20**: 2092-2099
- 50 **Nuyts J**, De Man B, Dupont P, Defrise M, Suetens P, Mortelmans L. Iterative reconstruction for helical CT: a simulation study. *Phys Med Biol* 1998; **43**: 729-737
- 51 **Sun Z**, Choo GH, Ng KH. Coronary CT angiography: current status and continuing challenges. *Br J Radiol* 2011; In press
- 52 **Donnino R**, Jacobs JE, Doshi JV, Hecht EM, Kim DC, Babb JS, Srichai MB. Dual-source versus single-source cardiac CT angiography: comparison of diagnostic image quality. *AJR Am J Roentgenol* 2009; **192**: 1051-1056
- 53 **Fang XM**, Chen HW, Hu XY, Bao J, Chen Y, Yang ZY, Buckley O, Wu XQ. Dual-source CT coronary angiography without heart rate or rhythm control in comparison with conventional coronary angiography. *Int J Cardiovasc Imaging* 2010; **26**: 323-331
- 54 **Ropers U**, Ropers D, Pflederer T, Anders K, Kuettner A, Stilianakis NI, Komatsu S, Kalender W, Bautz W, Daniel WG, Achenbach S. Influence of heart rate on the diagnostic accuracy of dual-source computed tomography coronary angiography. *J Am Coll Cardiol* 2007; **50**: 2393-2398
- 55 **Leber AW**, Johnson T, Becker A, von Ziegler F, Tittus J, Nikolaou K, Reiser M, Steinbeck G, Becker CR, Knez A. Diagnostic accuracy of dual-source multi-slice CT-coronary angiography in patients with an intermediate pretest likelihood for coronary artery disease. *Eur Heart J* 2007; **28**: 2354-2360
- 56 **de Graaf FR**, Schuijf JD, van Velzen JE, Kroft LJ, de Roos A, Reiber JH, Boersma E, Schalij MJ, Spanó F, Jukema JW, van der Wall EE, Bax JJ. Diagnostic accuracy of 320-row multidetector computed tomography coronary angiography in the non-invasive evaluation of significant coronary artery disease. *Eur Heart J* 2010; **31**: 1908-1915
- 57 **Pasricha SS**, Nandurkar D, Seneviratne SK, Cameron JD, Crossett M, Schneider-Kolsky ME, Troupis JM. Image quality of coronary 320-MDCT in patients with atrial fibrillation: initial experience. *AJR Am J Roentgenol* 2009; **193**: 1514-1521

S- Editor Cheng JX L- Editor Webster JR E- Editor Zheng XM