

World Journal of *Otorhinolaryngology*

World J Otorhinolaryngol 2019 April 27; 8(1): 1-11





EDITORIAL

- 1 CX3CR1 receptor as a potential therapeutic target in chronic rhinosinusitis and allergic rhinitis
El-Shazly A

CASE REPORT

- 4 Effect of intranasal stents (AlaxoLito, AlaxoLito Plus and AlaxoLito Xtreme) on the nasal airway: A case report
Zhang H, Kotecha B

Contents

World Journal of Otorhinolaryngology
Volume 8 Number 1 April 27, 2019

ABOUT COVER

Editor-in-Chief of *World Journal of Otorhinolaryngology*, Amr El-Shazly, MD, PhD, Professor, Department of Otorhinolaryngology, Liege University Hospital, Liege 4000, Belgium

AIMS AND SCOPE

World Journal of Otorhinolaryngology (*World J Otorhinolaryngol*, *WJO*, online ISSN 2218-6247, DOI: 10.5319) is a peer-reviewed open access academic journal that aims to guide clinical practice and improve diagnostic and therapeutic skills of clinicians.

The *WJO* covers topics concerning endoscopy, rhinology, pharyngology, laryngology, tracheo-esophagology, otology, tracheology, cancer, nasal symptomatology, congenital nasal diseases, inflammatory diseases of the external nose, rhinitis, allergic rhinitis, nasal polyps, nasal septal diseases, nasal bleeding, nasal or sinus foreign bodies, sinusitis, rhinogenic complications, diagnostic imaging, evidence-based medicine, epidemiology and nursing.

We encourage authors to submit their manuscripts to *WJO*. We will give priority to manuscripts that are supported by major national and international foundations and those that are of great basic and clinical significance.

INDEXING/ABSTRACTING

The *WJO* is now indexed in China National Knowledge Infrastructure (CNKI), China Science and Technology Journal Database (CSTJ), and Superstar Journals Database.

RESPONSIBLE EDITORS FOR THIS ISSUE

Responsible Electronic Editor: Yan-Liang Zhang Proofing Editorial Office Director: Ya-Juan Ma

NAME OF JOURNAL

World Journal of Otorhinolaryngology

ISSN

ISSN 2218-6247 (online)

LAUNCH DATE

December 28, 2011

FREQUENCY

Irregular

EDITORS-IN-CHIEF

Amr El-Shazly

EDITORIAL BOARD MEMBERS

<https://www.wjgnet.com/2218-6247/editorialboard.htm>

EDITORIAL OFFICE

Ya-Juan Ma, Director

PUBLICATION DATE

April 27, 2019

COPYRIGHT

© 2019 Baishideng Publishing Group Inc

INSTRUCTIONS TO AUTHORS

<https://www.wjgnet.com/bpg/gerinfo/204>

GUIDELINES FOR ETHICS DOCUMENTS

<https://www.wjgnet.com/bpg/GerInfo/287>

GUIDELINES FOR NON-NATIVE SPEAKERS OF ENGLISH

<https://www.wjgnet.com/bpg/gerinfo/240>

PUBLICATION MISCONDUCT

<https://www.wjgnet.com/bpg/gerinfo/208>

ARTICLE PROCESSING CHARGE

<https://www.wjgnet.com/bpg/gerinfo/242>

STEPS FOR SUBMITTING MANUSCRIPTS

<https://www.wjgnet.com/bpg/GerInfo/239>

ONLINE SUBMISSION

<https://www.f6publishing.com>

© 2019 Baishideng Publishing Group Inc. All rights reserved. 7041 Koll Center Parkway, Suite 160, Pleasanton, CA 94566, USA

E-mail: bpgoffice@wjgnet.com <https://www.wjgnet.com>



Effect of intranasal stents (AlaxoLito, AlaxoLito Plus and AlaxoLito Xtreme) on the nasal airway: A case report

Henry Zhang, Bhik Kotecha

ORCID number: Henry Zhang (0000-0003-3744-736X); Bhik Kotecha (0000-0001-7899-6707).

Author contributions: Zhang H and Kotecha B designed, investigated and wrote the manuscript.

Informed consent statement: The subject of this case report gave informed written consent for all procedures and investigations. He gave written consent for freedom of information for the publication.

Conflict-of-interest statement: There are no conflicts of interest declared by either author.

Open-Access: This article is an open-access article which was selected by an in-house editor and fully peer-reviewed by external reviewers. It is distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>

Manuscript source: Unsolicited Manuscript

Received: September 30, 2018

Peer-review started: October 2, 2018

First decision: December 5, 2018

Revised: March 4, 2019

Accepted: March 24, 2019

Article in press: March 25, 2019

Henry Zhang, Bhik Kotecha, Otolaryngology, Royal National Throat Nose and Ear Hospital, London WC1X 8DA, United Kingdom

Corresponding author: Henry Zhang, MBBS, Doctor, Otolaryngology, Royal National Throat Nose and Ear Hospital, 330 Grays Inn Rd, London WC1X 8DA, United Kingdom.

drhenryzhang@gmail.com

Telephone: +44-20-7921265304

Abstract

BACKGROUND

The study of intranasal stents on the nasal airway is limited in the medical literature. The authors aim to provide objective measurements on their effects on the nasal airway. The aim is to study the feasibility of three novel intranasal stenting devices, AlaxoLito, AlaxoLito Plus, and AlaxoLito Xtreme, as treatment for nasal obstruction.

CASE SUMMARY

A 58-year-old man, who had right-sided nasal obstruction, used stents during sporting activities intermittently for four years and subsequently in addition to intermittent sports use regularly for sleep for another two years. Magnetic resonance imaging (MRI) of the nasal passages and rhinomanometric measurements were taken with and without stents *in situ*. The stents tested are all braided from thin nitinol wires. The AlaxoLito Nasal Stent has a length of 35 mm. The AlaxoLito Plus and AlaxoLito Xtreme Nasal Stents have a length of 60 mm. Both have a diameter of about 10 mm in unloaded state and comprise a widened, ball-shaped section (which is positioned at the nasal alar) of about 11 and 14 mm, respectively. Rhinomanometric nasal airflow after application of the stents improved 1.11, 1.23, and 1.38 fold, respectively, with application of the AlaxoLito, AlaxoLito Plus and AlaxoLito Xtreme stents. MRI showed that after application of the stents, the nasal passage increased in diameter.

CONCLUSION

Intranasal stenting shows improvement in nasal airflow. Intermittent and regular long-term use had been shown to be safe, with no discomfort and no side effects.

Key words: Intranasal stent; Airway; Obstructive sleep apnoea; Rhinomanometry; Magnetic resonance imaging

©The Author(s) 2019. Published by Baishideng Publishing Group Inc. All rights reserved.

Published online: April 27, 2019

P-Reviewer: El-Shazly A

S-Editor: Cui LJ

L-Editor: A

E-Editor: Zhang YL



Core tip: Intranasal stents have a number of applications in the clinical setting, but no current study existed to determine objective measurements. This case shows objective data on rhinomanometry and imaging measurements, to determine long term effects of three types of intranasal stents on the nasal airway.

Citation: Zhang H, Kotecha B. Effect of intranasal stents (AlaxoLito, AlaxoLito Plus and AlaxoLito Xtreme) on the nasal airway: A case report. *World J Otorhinolaryngol* 2019; 8(1): 4-11

URL: <https://www.wjgnet.com/2218-6247/full/v8/i1/4.htm>

DOI: <https://dx.doi.org/10.5319/wjo.v8.i1.4>

INTRODUCTION

A variety of devices exist to treat nasal obstruction, ranging from nasal valve dilator devices (Max-Air Nose Cones) to external devices (Breathe Right nasal strip), and these have been shown to improve nasal airflow^[1].

Nasopharyngeal stenting, which splints open the velopharynx, has been used to treat sleep disordered breathing^[2], and plays a role in surgical planning for Obstructive Sleep Apnoea (OSA)^[3,4]. Previous studies of nasopharyngeal stenting devices have shown reduced snoring and OSA as evidenced by an improvement in audiometric data and the apnoea-hypopnoea index during sleep^[5]. These devices however, are not designed to splint the nasal airway and to improve the nasal breathing.

Fluid dynamics studies of nasal airflow have shown that a larger proportion of airflow is through the middle nasal passage, drawing conclusions that ensuring optimal airflow through this area may provide better laminar flow through the nasal passage^[6,7]. This led to the development of intra-nasal stenting devices.

This article presents for the first time, a single case report with patient experience using intranasal stents (Figure 1), the AlaxoLito, AlaxoLito Plus and AlaxoLito Xtreme Nasal Stents. The authors describe symptomatic treatment including changes in objective measurements, imaging and rhinomanometric data.

MEDLINE, Scopus, and the Cochrane Library were searched from inception to September 2018, for treatment of nasal obstruction, rhinitis, snoring with a nasal stenting. Terms searched included the various combinations of nasal obstruction and nasal trumpet, stent, obturator, airway or tube. There were no studies in Medicine describing the use in nasal obstruction or rhinitis.

CASE PRESENTATION

A 58 years old healthy male volunteer enrolled into clinical testing for the use of three intranasal stents (AlaxoLito, AlaxoLito Plus and AlaxoLito Xtreme). The patient had a previous tonsillectomy, and no other significant past medical history. Symptomatically he had mild right sided nasal obstruction. On clinical examination, he had a mildly deviated nasal septum to the right. Anterior rhinomanometry (Flowhandy ZAN 100 USB, ZAN Meßgeräte GmbH, Oberthulba, Germany; flow at 150 mbar pressure difference) showed a baseline of 370L/min on the right and 380L/min on the left nasal passage.

FINAL DIAGNOSIS

Right nasal septal deviation associated with mild non-allergic rhinitis.

TREATMENT

The subject applied the intranasal stents for a total of 6 years, starting in 2012. All stents always have been inserted into the middle nasal passage. The application of these stents was performed by the subject themselves. The stents have an insertion tube which facilitates insertion, which is removed after the stents are in a comfortable




Figure 1 AlaxoLito (A) and AlaxoLito Plus (C-E) Nasal Stents with insertion tube. Insertion method: B and C: The stent is placed inside the insertion tube, which allows for easy insertion; D: Once the insertion tube is placed into the nasopharynx; E: It is withdrawn, allowing the stent to open inside the nasal cavity.

position. For the first four years, the subject applied the stents only during exercise and sporting activities, AlaxoLito Plus on the left nasal passage and AlaxoLito on the right nasal passage, due to the mild septal deviation. The AlaxoLito positioned just in front of the septum deviation. After regular nightly use of these stents for determination of the effects during sleep from May 2016 on, after two weeks the subject noticed he could insert the longer AlaxoLito Plus stent also on the right side, initially only in the morning, and after about two months around the clock, and without pain. Since then, the subject used the AlaxoLito Plus on both sides. In May 2018 the subject started using a new stronger variant of the AlaxoLito Plus (the AlaxoLito Xtreme) on both sides.

OUTCOME AND FOLLOW-UP

Effects on rhinomanometry

Subjectively, the patient experienced an improvement in nasal airflow with application of the stent, especially during physical activity, but also during sleep. He also reported an improvement in nasal airflow when applied during an episode of upper respiratory tract infection.

Nasal rhinomanometry was performed without and with the application of the three stent types. Mean measurements were taken to account for the nasal cycle. The results are shown in [Figure 2](#). An increase in nasal airflow from 751 L/min to 832

L/min was observed with AlaxoLito stent. A further increase to 925 L/min and 1036 L/min, respectively, was measured with the application of the AlaxoLito Plus and AlaxoLito Xtreme stents. The subject reported that inspiration was hampered by the olive of the rhinomanometry system starting with the AlaxoLito Plus and the more for the AlaxoLito Xtreme, due to a reduced diameter of the possible air inflow into the nasal valve. Thus, the real values and factors of increase seem to be higher than measured.

Ten horizontal and vertical points of measurements of the nasal airway were taken in the coronal plane, and the mean of these calculated. Using this method, an increase in the diameter of the middle nasal airway of 1mm was seen from 2015 to 2017 without the nasal stents *in situ*, and of 2.3 mm with the AlaxoLito Plus *in situ*. The strongest widening occurred at the position of the septum deviation in the right nasal passage with 4.1 mm. With the AlaxoLito Xtreme additional widening and an almost round shape of the air canal was observed.

Imaging studies

Multiple magnetic resonance imaging (MRI) studies were performed. **Figure 3** shows a comparison of coronal MRI of the nasal passages prior to regular nightly use of the stents (December 2015) and with about 1 year post regular AlaxoLito Plus stent application (June 2017). A clear improvement by widening of the nasal airway post regular stent usage is seen on the MRI, especially on the right side, as evidenced by the increase in the black area of the middle nasal passage representing the air canal provided by the stent (red arrows). A comparative MRI was performed with the stents *in situ*. The yellow arrows show a compression effect they have on the middle turbinates. Further, a certain reduction by rounding of the probably cartilaginous structure of the septum deviation is observable (red arrows).

Summary comparisons of the middle nasal airway without and with stents are shown in **Figure 4**. Without stents, the middle nasal airway has improved on both sides from December 2015 to June 2017, and again in September 2018. With the stents *in situ*, a clear improvement by increase of the cross-sectional area in the middle nasal airway is shown, with the shape having become more or less round due to the application of the AlaxoLito Xtreme.

In **Figure 5** axial and sagittal planes in MRI are shown for September 2018 with the AlaxoLito Xtreme stents *in situ*. It is evident that the stents create a rather consistent airway of about 6.5 mm diameter.

Table 1 lists the diameters of the right nasal passage at ten distinct points along its entire length. The values measured demonstrate (1) a clear increase in width and height of the nasal passage over its entire length without stent *in situ* (comparison 2015-12 to 2017-05); (2) an additional clear increase in width and height of the nasal passage over its entire lengths when comparing without and with stent *in situ* (2017-05); (3) the narrow space at the position of the local septum deviation (positions number 6 to 8) has been significantly widened (2017-05) without the stent *in situ* so that a more uniform shape of the nasal passage results. Insertion of the AlaxoLito Plus leads to a rather consistent tubular shape of the nasal passage. A similar result in terms of widening of the nasal passage has been obtained for the left side (data not shown).

DISCUSSION

Although nasopharyngeal airway tubes have been used by both anaesthesiologists and otolaryngologists to treat upper airway obstruction at the level of the velopharynx, the treatment of nasal airway obstruction using stenting devices has limited prior research.

The AlaxoLito, AlaxoLito Plus and AlaxoLito Xtreme stents are braided from thin nitinol wires, a memory metal. For the patient in this case report, all stents provided a subjective improvement in nasal breathing, with stepwise increasing effect from AlaxoLito to AlaxoLito Plus to AlaxoLito Xtreme.

Rhinomanometric measurement results improved stepwise with application of AlaxoLito to AlaxoLito Plus to AlaxoLito Xtreme. The longer AlaxoLito Plus stent was initially not insertable on the right side, where the patient had mild obstruction secondary to septal deviation, as the 4 mm diameter insertion tube could not pass the narrow position at the septum deviation, instead causing pain on the septum mucosa. The intermittent and irregular use of the shorter AlaxoLito mainly during sports, each time for some hours, over four years did not change this situation. However, after regular nightly use of the AlaxoLito stent for two weeks, the patient was able to insert the longer AlaxoLito Plus on the right side without pain and discomfort, indicating




Figure 2 Rhinomanometric data without and with stents.

there may be cumulative effects on opening the nasal airway by regular nightly application for about 8 h per night. Initially this effect lasted for 2 to 3 h only. After about 2 mo this effect lasted around the clock so that the longer AlaxoLito Plus could be inserted also in the evening before sleep and was used regularly since then.

The use of the stronger version (the AlaxoLito Xtreme) again within 2 to 3 mo led to a further widening of the middle nasal airway at the position of the septum deviation. Although there was no resolution of the septal deviation, the nasal airway on the narrow side improved with measurements on MRI. This suggests that a period of about 2-3 mo of use is necessary to achieve the permanent widening effects which are visible in the MRI. All MRI studies have been performed in the evening, that means at least 10 h or more after removal of the stents from the nose in the morning.

An increase of the diameter of the middle nasal passage of 1 mm without stents and 2.3 mm with stents *in situ* was observed after one year of regular use of the AlaxoLito Plus stent spanning across this septum deviation position. 5.5 to 6.5 mm hydraulic diameter is considered to be the normal range of the nasal passage for good nasal breathing^[8]. With the stents an almost ideal tube has been reached, so that the hydraulic tube diameter values should be directly applicable for comparison.

This is the first case report to examine the effects of intranasal stents in the treatment of nasal obstruction, using rhinomanometric data and MRI. Clear improvement in the nasal airway is shown on MRI, both with the stent *in situ*, and after regular nightly use of the stent for one year in the first step (AlaxoLito Plus) and 2-3 mo in the second step (AlaxoLito Xtreme). Further imaging and rhinomanometric data would be helpful after the patient ceases stent use, to analyse whether the stents have a longterm lasting effect on the nasal airway.

Although the patient reported subjective improvement in sleep quality, however more objective data would be needed to quantify this. Occasional cell phone app recordings (SnoreLab) indicated almost complete abolishment of occasional weak snoring finally with the AlaxoLito Xtreme. However, the authors did not systematically study the effect of these intranasal stents on snoring or sleep apnoea, and further studies would be warranted to study their effects on sleep.

Table 1 Diameters of the right nasal passage at ten distinct points along its entire length (from anterior to posterior)

2015-12 without stent <i>in situ</i>			2017-05 without stent <i>in situ</i>			2017-05 with AlaxoLito Plus		
Width(mm)	Height(mm)	Sum(mm)	Width(mm)	Height(mm)	Sum(mm)	Width(mm)	Height(mm)	Sum(mm)
4.54	3.80	8.34	4.78	4.55	9.33	5.63	6.04	11.67
3.53	4.53	8.06	3.97	5.51	9.48	5.66	6.25	11.91
2.01	4.88	6.89	3.18	5.73	8.91	4.62	5.90	10.52
2.27	4.21	6.48	3.59	6.06	9.65	4.61	5.88	10.49
3.50	3.53	7.03	3.77	5.06	8.83	5.49	6.64	12.13
3.37	1.79	5.16	4.72	3.19	7.91	6.64	6.07	12.71
3.22	2.56	5.78	3.61	3.53	7.14	6.50	5.67	12.17
2.70	4.79	7.49	3.95	6.53	10.48	6.35	6.89	13.24
3.29	4.93	8.22	5.03	5.40	10.43	6.51	6.76	13.27
5.65	4.00	9.65	7.27	5.16	12.43	7.58	7.15	14.73




Figure 3 Comparison of coronal magnetic resonance imaging of nasal passages prior to and 1 year post regular stent use. A: Magnetic resonance imaging (MRI) from December 2015 prior to regular use of stents, without stents; B and C: MRI from June 2017 one year post regular use of AlaxoLito Plus stent (B: without stents *in situ*; C: with AlaxoLito Plus *in situ*).




Figure 4 Coronal magnetic resonance imaging showing measurements of middle nasal passage without stents (A-C) and with stents *in situ* (D-F). A and D: Magnetic resonance imaging (MRI) from December 2015 prior to regular use of stents; B and E: MRI from June 2017 one year post regular use of AlaxoLito Plus; C and F: MRI from September 2018, 3.5 mo post regular use of AlaxoLito Xtreme.




Figure 5 Axial (A) and sagittal (B) magnetic resonance imaging (from September 2018) showing measurements of middle nasal passage with AlaxoLito Xtreme stents *in situ*.

REFERENCES

- 1 **Raudenbush B.** Stenting the nasal airway for maximizing inspiratory airflow: internal Max-Air Nose Cones versus external Breathe Right strip. *Am J Rhinol Allergy* 2011; **25**: 249-251 [PMID: [21819762](#) DOI: [10.2500/ajra.2011.25.3621](#)]
- 2 **Camacho M,** Chang ET, Fernandez-Salvador C, Capasso R. Treatment of Snoring with a Nasopharyngeal Airway Tube. *Case Rep Med* 2016; **2016**: 3628716 [PMID: [27795710](#) DOI: [10.1155/2016/3628716](#)]
- 3 **Zhang H,** Kotecha B. Using Nasopharyngeal Stenting Devices as a Novel Way of Surgical Planning for Obstructive Sleep Apnea. *J Clin Sleep Med* 2018; **14**: 491 [PMID: [29458707](#) DOI: [10.5664/jcsm.7016](#)]
- 4 **Powell HR,** Watson NA, Kotecha BT. Pilot study assessing the efficacy of a novel treatment for sleep-related breathing disorders in patients undergoing sleep nasendoscopy: Our Experience. *Clin Otolaryngol* 2014; **39**: 190-194 [PMID: [24801561](#) DOI: [10.1111/coa.12253](#)]
- 5 **Traxdorf M,** Hartl M, Angerer F, Bohr C, Grundtner P, Iro H. A novel nasopharyngeal stent for the treatment of obstructive sleep apnea: a case series of nasopharyngeal stenting versus continuous positive airway pressure. *Eur Arch Otorhinolaryngol* 2016; **273**: 1307-1312 [PMID: [26525882](#) DOI: [10.1007/s00405-015-3815-2](#)]
- 6 **Keyhani K,** Scherer PW, Mozell MM. Numerical simulation of airflow in the human nasal cavity. *J Biomech Eng* 1995; **117**: 429-441 [PMID: [8748525](#) DOI: [10.1109/1EMBS.2005.1615757](#)]
- 7 **Castro Ruiz P,** Castro Ruiz F, Costas López A, Cenjor Español C. Computational fluid dynamics simulations of the airflow in the human nasal cavity. *Acta Otorrinolaringol Esp* 2005; **56**: 403-410 [PMID: [16353786](#)]
- 8 **Mlynski G.** New Measurement Methods in the Diagnostic of Nasal Obstruction. *Nasal Physiol & Pathophysiol Nasal Disorders* 2013; 361-388 [DOI: [10.1007/978-3-642-37250-6_27](#)]



Published By Baishideng Publishing Group Inc
7041 Koll Center Parkway, Suite 160, Pleasanton, CA 94566, USA
Telephone: +1-925-223-8242
Fax: +1-925-223-8243
E-mail: bpgoffice@wjgnet.com
Help Desk: <https://www.f6publishing.com/helpdesk>
<https://www.wjgnet.com>

