

79514_Auto_Edited.docx

Name of Journal: *World Journal of Diabetes*

Manuscript NO: 79514

Manuscript Type: MINIREVIEWS

The role of artificial intelligence technology ² in the care of diabetic foot ulcers: the past, the present, and the future

Artificial intelligence technology ² in the care of diabetic foot ulcers

Joseph M Pappachan, Bill Cassidy, Cornelius James Fernandez, Vishnu Chandrabalan,
Moi Hoon Yap

Abstract

3
Foot ulcers are common complications of diabetes mellitus and substantially increase the morbidity and mortality from the disease. Wound care by regular monitoring of the progress of healing with clinical review of the ulcers, changes in dressing, appropriate antibiotic therapy for infection and proper offloading of the ulcer are the cornerstones of the management of foot ulcers. Assessing the progress of foot ulcers can be a challenge for the clinician and patient because of logistic issues such as regular attendance in the clinic. Foot clinics are often busy and because of manpower issues, ulcer reviews can be delayed with detrimental effects of the healing from lack of appropriate and timely changes in management. Wound photographs have been historically useful to assess the progress of diabetic foot ulcers over the past few decades. Mobile phones with digital cameras have recently revolutionized the capture of foot ulcer photographs. Patients can send ulcer photographs to diabetes care professionals electronically for remote monitoring, largely avoiding logistics of patient transport to clinics with a reduction on clinic pressures. Artificial intelligence-based technologies have been developed in recent years to improve this remote monitoring of diabetic foot ulcers with the use of mobile apps. This is expected to make a huge impact on diabetic foot ulcer care with further research and development of more accurate and scientific technologies in future. This clinical update review is to compile evidence on this hot topic to empower clinicians with the latest developments in the field.

Key Words: Diabetic foot ulcers (DFU); Photographic monitoring; Artificial intelligence technology; Digital photography; Mobile app; COVID-19 pandemic

Pappachan JM, Cassidy B, Fernandez CJ, Chandrabalan V, Yap MH. The role of artificial intelligence technology in the care of diabetic foot ulcers: the past, the present, and the future. *World J Diabetes* 2022; In press

Core Tip: Diabetic foot clinics faced major challenges during the COVID-19 pandemic due to lockdowns and social distancing measures as a significant proportion of patients were unable to physically attend the clinics. This situation boosted up the attempts for transition of face-to-face foot clinics to virtual clinics as in many other types of medical care during the pandemic. Monitoring of diabetic foot ulcers (DFU) by digital photographic technology and mobile phone-based photography are revolutionising this area of clinical care in the recent years and mobile apps are expected to accelerate this progress. This review is to update the past, the present and the future of artificial intelligence technology in the care of DFU.

INTRODUCTION

Foot ulceration caused by micro- and/or macrovascular disease is ³ a common complication of diabetes mellitus (DM). The global prevalence of diabetic foot ulcers (DFU) at any point in time is estimated to be 6.3%, ^[1] with 25% of DM patients developing a DFU in their lifetime ^[2]. DFU significantly increases morbidity (including lower extremity amputations [LEA]) and mortality among sufferers, and the care of DFU poses very high healthcare expenditure all over the world. Approximately 50% of DFU are associated with infections ^[3], and $\approx 20\%$ of moderate to severe DFU infections may lead on to minor or major amputations ^[4]. DFU-related LEA is associated with a 10-year survival rate as low as 24% which is lower than that of several forms of cancers ^[5]. These alarming figures give us a broad outline of the health-related risks posed by DFUs.

Appropriate care of DFU for ensuring rapid cure to avoid complications involve regular monitoring of the progress of healing with periodic review of the ulcers, frequent dressing changes, antibiotic therapy for infection control, optimal control of DM, and adequate offloading of the ulcers to avoid ongoing damage from pressure on the wound area. Diabetic foot clinics usually provide comprehensive DFU care by multidisciplinary specialist teams involving podiatrists, orthotists (who make appropriate footwear/devices for offloading the DFU), diabetes specialist nurses, and

diabetologists. However, insufficient manpower, long clinic waiting lists, and logistic issues with patient transport to the clinics can all pose problems for timely review of patients with DFU in the foot clinics. Manpower shortage resulting in longer waiting periods prior to review by footcare team had been a major issue in ulcer care in diabetes foot services across the world in recent years [6, 7]. Delays in presentation for foot ulcer review had been identified as an important reason for wound healing [8]. Complications of foot ulcers including amputations can be the devastating sequel of such delays in review by footcare team.

The COVID-19 pandemic posed major challenges in DFU across the globe due the above reasons [9-11]. Delays in ulcer reviews may be associated with detrimental effects on the DFU healing that may even lead on to amputations [10,11]. Telephone and video clinics during the pandemic period helped clinicians to address some of the issues related to inability of patients to attend medical clinics during the COVID-19 Lock-down period. Attempts for the development of artificial intelligence (AI) algorithms by the scientific fraternity to enhance patient care through virtual clinics attained greater momentum in relation to the pandemic [12,13].

Monitoring the progress of DFU by comparison of serial ulcer photographs during the follow up in foot clinics is practiced by many diabetologists over the past 2-3 decades. Refinement of the photographic methods by sophisticated digital technology and mobile phone cameras have revolutionized DFU care in the recent years [14-16]. Some patients often use mobile phone cameras to capture their foot ulcer photographs for self-monitoring wound progress and to help clinicians to understand the previous status of their wound during clinic review.

However, much more multidisciplinary scientific research input and output are necessary in this area. Further refinement of this rapidly advancing digital technology for its optimal use in day-to-day clinical practice could result in better care of patients with DFU. This review attempts to gather the up-to-date evidence to summarise the past, present, and future dimensions of AI algorithms to empower clinicians across the globe to appropriately utilize digital technology for DFU care.

Photographic monitoring of DFU progress

Diabetes care professionals often see several cases of DFU in their day-to-day clinical practice and often forget the previous ulcer grade, character and even the site during subsequent visits, weeks later when followed up in the foot clinic. Review of the previous photographs during subsequent follow up visits should provide clinicians good clinical assessment of the progress of the ulcer and help prognostication [17]. Photographic monitoring also enhances appropriate continuity of care by different clinicians running the diabetic foot services during their review of the same case.

Important issues which can arise while comparing the photographs are the differences in the distance from which photographs are captured (can be largely avoided by placing a measuring tape on the ulcer while photographed), differences in the illumination of pictures from brightness of background light when the photographs were taken, and the likelihood of low-quality images without adequate focussing of cameras by individuals taking the picture. Photographs are usually taken without flashlight to avoid undue illumination that can reduce the picture quality. Clinical evidence suggests that prediction of ulcer healing is possible by regular photographic monitoring [17]. *Figure 1 A, B, C & D* shows the photographic monitoring of progress of a DFU at various stages.

History of digital archiving of the DFU photographs

Use of digital cameras for photography became a common practice since early 1980s after the invention of digital photographic technology in 1975. Digital archiving of foot ulcer photographs in computer databases was a great leap forward in the technology for the monitoring of DFU. Print outs of foot ulcer photographs are still used in the remote settings where internet and computers are not freely available in clinical practice. This is in fact more expensive (costs incurred with colour printing and use of good quality photographic paper) and cumbersome in the modern era.

Fading out of the colour print outs over time makes the situation worse regarding DFU monitoring in this method. Review of the serial images in the computer database of

digital photographs makes the work of diabetes foot care professionals much easier wherever the facility is available [18]. Lack of degradation of image quality over time as in printed photographs is another great advantage of digital archiving of DFU images. Mobile phones became popular for telephone communications globally in the early 2000s and newer versions of these devices with cameras and video recording facilities came into the market within next few years. With rapid advancements in technology, smartphone devices using high resolution cameras are now widely available and became an integral part of modern life in the 21st century. Diabetes care professionals soon started using mobile phone cameras and video facilities for monitoring and management of DFUs [19-21]. Some patients themselves were using mobile phones for picturising their DFUs prior to their foot clinic visits so that they could show the status of their ulcers earlier. Although the initial studies on use of mobile phones for monitoring of DFU were not very promising,²² subsequent studies show encouraging results [23-25].

Integrating DFU photographs with computer-based digital technology for diagnostics

Artificial Intelligence (AI) and its applications have been utilised in various branches of modern science and technology including the medical field over the past few decades to improve physical and intellectual human work output. Such technologies can also be utilised for monitoring the progress DFU care. DFU datasets were utilised for training and testing the machine learning processes. Collection of such datasets supports the ongoing research for DFU academic AI challenges, such as those hosted by the Medical Image Computing and Computer Assisted Interventions (MICCAI) conferences [26,27]. These challenges are used to promote and advance ongoing research and to increase exposure within the associated academic fields.

Machine learning algorithms have been found to be very useful in detecting DFUs with high accuracy rates in previous studies [18-30]. These algorithms are developed by using large datasets of images captured from the foot clinics. The development of computer-aided DFU diagnostic algorithms involves multiple stages such as pre-processing,

feature extraction, detection, classification, and segmentation of DFU wounds [30]. These tasks can be challenging in real-world settings due to the low-quality of images from inadequate focussing, motion artifacts, inadequate lighting, and backlight, deformities in the foot/toes, size and shape of the ulcers (very small, very large and curved ulcers), and newly formed or early ulcers which are easily missed during capture of photographs from foot clinics [16].

The first and foremost step in the development of machine learning algorithms is detection of DFU from foot photographs. This task has been successfully done by previous researchers as mentioned above [28-30]. Further refinement of machine learning is currently being undertaken by incorporating the classification systems for DFU, as in the real-world settings, to promote AI-based diagnostics and prognostication. Several manual DFU classification systems are currently used by foot care professionals such as Wagner, University of Texas, and ⁵ SINBAD (Site, Ischaemia, Neuropathy, Bacterial infection, Area, and Depth) for DFU monitoring and management [30]. These manual approaches may be able to benefit from the automated processes afforded by AI.

Incorporation of such complex features in the AI-technology for day-to-day clinical practice to enhance diagnosis and prognostication can be challenging. These challenges include: (1) the significant time-burden involved in the DFU image data collection and appropriate labelling; (2) the inter- and intra-class variations depending on differences in the classification of DFU; (3) lack of standardization of DFU datasets (caused by camera distance from foot, image orientation and lighting conditions); (4) the differences in ethnicity, age, sex, and foot size of the patients [30]. Development of deep learning AI algorithms requires large-scale datasets for automated DFU analysis to reproduce comparable results at par with those from experts. Researchers currently working in isolation may not achieve reproducible research outputs. Large DFU datasets used for training and validation by multiple professionals from different institutions across the globe should help to refine these pitfalls in machine learning algorithms for DFU classification and diagnosis. To enable innovation from clinicians and researchers, ¹ Yap *et al* proposed the diabetic foot ulcer challenges by providing the

¹ publicly available datasets, for comprehensive evaluation of object detection frameworks on DFU detection using convolutional neural networks trained on the DFUC2020 dataset [31,32]. Examples of manually delineated DFU photographs from the training set are shown in *Figure 2* and *Figure 3*. Morphological classification of DFU into different types (such as infection, ischemia, both, and none) is done in these datasets to enable machine learning.

Serial demarcation of ulcer photographs to assess progress of DFU healing

At present monitoring the healing process of DFU is largely completed by regular patient follow-up visits in the multidisciplinary foot clinics [33-35]. Since 2020, the COVID-19 pandemic has made a huge impact on DFU care across the world due to the issues related to lockdowns and social distancing measures enforced by governments to curtail the human devastation from the global health emergency. Foot clinic attendance [36] and hospitalisations from DFU [37] were significantly reduced during the pandemic. Although worsening of DFU were comparatively less owing to reduced outdoor human activity [38], many patients were presented to clinicians late, increasing the risk of complications from DFU [37,39]. The pandemic reinforced the urgent necessity to develop AI-based wound care algorithms for remote monitoring of DFU.

To enable development of machine learning algorithms, serial demarcation of DFU ¹ is important to assess the progress of healing process along with the wound classification based on one of the above-mentioned standard systems. Therefore, characterisation of DFU requires demarcation of wounds at various stages in the datasets by experts in the field after which the data can be used for training, validation, and testing. This task is highly labour intensive and requires huge DFU training and validation datasets.

Photographs acquired by mobile phone cameras for DFU monitoring

New generation mobile phones can capture high resolution DFU photographs at par with commercial digital cameras. Many patients already use mobile phone photography for self-monitoring their DFU [14,22]. Some patients serially document their wound progress through these pictures and meticulously bring them to foot clinics

during their follow-up visits for DFU care. Occasionally, patients send their photographs to footcare professionals electronically or as email attachments. Though safety issues related to safe transfer of such clinical information without breach of confidentiality can be challenging at present, wider use of mobile phone photography for self-monitoring of DFU progress might revolutionise future foot care.

Development of mobile apps for DFU monitoring

Various mobile phone apps are currently used worldwide in many domains of daily activities to improve the quality human life. Such mobile apps for DFU monitoring and care are under development. Cassidy *et al* developed the first mobile app capable of accurate DFU detection using AI and cloud-based technologies ^[40,41]. This system was tested in a 6-month clinical evaluation at two UK National Health Service hospital sites (Lancashire Teaching Hospitals and Salford Royal Hospital) and is currently being further developed to improve functionality and accuracy. Additional app features, such as automated classification of DFU wound pathology^[42] and automated delineation of wound / peri-wound regions are also being investigated ^[43] to provide a more clinically relevant system.

These new technological advancements are expected to revolutionise remote wound care by enabling patients to self-monitor their DFUs and contacting clinicians when they find any concern or deterioration of their disease. This would also enhance flexibility in the functioning of foot clinics by reviewing and triaging the most appropriate DFU to be seen in the clinics by remote monitoring of patients.

Prediction of DFU progress by integration of clinical and biochemical parameters and ulcer photographs

Although regular wound care by dressings, appropriate antibiotic therapy for infections, offloading the ulcers to relieve pressure-related delay in healing and revascularisation of ischemic foot are the cornerstones in management of DFU, several other clinical (comorbid conditions such as renal disease, heart failure, and immunosuppressed states) and biochemical parameters (such as hyperglycaemia, anemia, and high HbA1c) may impact DFU healing process ^[44-49]. Integration of these

clinical and biochemical parameters into machine learning algorithms should help us to develop prediction models using AI technology.

We note however that development of such algorithms is much more labour-intensive as demarcation exercises to develop deep learning models require larger datasets with clinical and biochemical parameters of individual patients incorporated within the neural network. However, appropriate use of computer technology integrated with digital applications **can help to reduce the physical burden on** researchers **in** developing such models.

Future perspectives

Recent research has investigated the utilisation of patient data in the training of deep neural networks in various medical imaging domains ^[50]. These studies indicate that machine learning models trained on patient data can be used to boost the performance of convolutional neural networks trained on wound / lesion images. Research of this nature is ongoing and represents a way to incorporate a more integrated method of wound analysis that considers multiple data points. Development of mobile apps integrating such new advancements in technology is expected to revolutionise the global scenario of DFU care in the near future.

CONCLUSION

Digital applications in the daily management of DFU have evolved rapidly in recent years to a level of remote diagnosis and monitoring of wounds in the community settings. The COVID-19 pandemic has accelerated research and development of such innovative technological applications in the past two years. Photographic monitoring of foot ulcers was practiced in many centres across the world in the past few decades providing DFU care. Invention of digital photographic technology in 1975 further boosted DFU care because of the ease of electronic archiving of ulcer pictures during clinical follow-up. Photography using mobile phone cameras became a huge leap forward in this direction in recent years empowering patients and clinicians to further improve the DFU care.

Artificial intelligence-based digital algorithms are currently being developed rapidly through collaborative global effort between AI experts and clinical teams. Mobile camera-based digital technologic applications are under development to enhance remote diagnosis, monitoring, and follow-up care of DFU. Prediction models of wound healing are also under development now making use of linking the ulcer characteristics DFU pictures to clinical and laboratory parameters of diabetic patients. These collaborative efforts between clinicians and computer scientists across the world should revolutionise such discoveries to empower diabetic foot patients to self-monitor and manage their DFUs to a greater extent.

3%

SIMILARITY INDEX

PRIMARY SOURCES

1

Moi Hoon Yap, Ryo Hachiuma, Azadeh Alavi, Raphael Brüngel et al. "Deep Learning in Diabetic Foot Ulcers Detection: A Comprehensive Evaluation", Computers in Biology and Medicine, 2021

Crossref

37 words — 1%

2

P. Sheehan, P. Jones, A. Caselli, J. M. Giurini, A. Veves. "Percent Change in Wound Area of Diabetic Foot Ulcers Over a 4-Week Period Is a Robust Predictor of Complete Healing in a 12-Week Prospective Trial", Diabetes Care, 2003

Crossref

21 words — 1%

3

Judith O'Brien, Alex Ward, Maribel Salas. "Managing diabetic foot ulcers: economic consequences in the USA", Expert Review of Pharmacoeconomics & Outcomes Research, 2014

Crossref

17 words — 1%

4

research.cs.queensu.ca

Internet

10 words — < 1%

5

www.ncbi.nlm.nih.gov

Internet

9 words — < 1%

