

89376_Auto_Edited.docx

Name of Journal: *World Journal of Orthopedics*

Manuscript NO: 89376

Manuscript Type: ORIGINAL ARTICLE

Observational Study

A single-center experience with Knee+™ augmented reality navigation system in primary total knee arthroplasty

Knee+™ augmented reality navigation system experience

Evangelos Sakellariou, Panagiotis Alevrogiannis, Fani Alevrogianni, Athanasios Galanis, Michail Vavourakis, Panagiotis Karampinas, Panagiotis Gavriil, John Vlamis, Stavros Alevrogiannis

Abstract

BACKGROUND

Computer-assisted systems obtained an increased interest in orthopaedic surgery over the last years, as they enhance precision compared to conventional hardware. The expansion of computer assistance is evolving with the employment of augmented reality. Yet, the accuracy of augmented reality navigation systems has not been determined.

AIM

This study aims to examine the accuracy of component alignment and restoration of the affected limb's mechanical axis in primary TKA, utilizing an augmented reality navigation system and to assess whether such systems are conspicuously fruitful for an accomplished knee surgeon

METHODS

From May 2021 to December 2021, 30 patients, 25 women and five men, underwent a primary unilateral total knee arthroplasty. Revision cases were excluded. A preoperative radiographic procedure was performed to evaluate the limb's axial alignment. All patients were operated on by the same team, without a tourniquet, utilizing three distinct prostheses with the assistance of the Knee+™ augmented reality navigation system in every operation. Postoperatively, the same radiographic exam protocol was executed to evaluate the implants' position, orientation and coronal plane alignment. We recorded measurements in 3 stages regarding femoral varus and flexion, tibial varus and posterior slope. Firstly, the expected values from the AR system were documented. Then we calculated the same values after each cut and finally, the same measurements were recorded radiologically after the operations. Concerning statistical analysis, Lin's concordance correlation coefficient was estimated, while Wilcoxon Signed Rank Test was performed when needed.

RESULTS

A statistically significant difference was observed regarding mean expected values and radiographic measurements for femoral flexion measurements only (Z score = 2.67, p -value = 0.01). Nonetheless, this difference was statistically significantly lower than 1 degree (Z score = -4.21, p -value < 0.01). In terms of discrepancies in the calculations of expected values and controlled measurements, a statistically significant difference between tibial varus values was detected (Z score = -2.33, p -value = 0.02), which was also statistically significantly lower than 1 degree (Z score = -4.99, p -value < 0.01).

CONCLUSION

The results indicate satisfactory postoperative coronal alignment without outliers across all three different implants utilized. Augmented reality navigation systems can bolster orthopaedic surgeons' accuracy in achieving precise axial alignment. However, further research is required to further evaluate their efficacy and potential.

Key Words: augmented reality; orthopedics; total knee arthroplasty; robotics; knee; navigation

Sakellariou E, Alevrogiannis P, Alevrogianni F, Galanis A, Vavourakis M, Karampinas P, Gavriil P, Vlamis J, Alevrogiannis S. A single-center experience with Knee+™ augmented reality navigation system in primary total knee arthroplasty. *World J Orthop* 2024; In press

Core Tip: Augmented reality navigation systems can bolster orthopaedic surgeons' accuracy in achieving precise axial alignment.

INTRODUCTION

Total Knee Arthroplasty (TKA) is indubitably one of the most common orthopaedic procedures ordinarily performed for knee joint osteoarthritis. In the USA, over 600,000 TKAs/year are carried out, with the numbers soaring annually [1].

The preponderant goals of TKA are pain relief, joint function restoration and prostheses' longevity. TKA's success and artificial joint longevity are achieved with the restoration of limb alignment. Malalignment is recurrently associated with copious long-term complications, such as tibiofemoral and patellofemoral instability, joint stiffness, patellar fractures, increased polyethylene wear and implant loosening [2,3]. Diligent comprehension of the fundamentals and the employment of precise instrumentation are pivotal for executing reproductively a well-aligned TKA.

Computer-assisted systems acquired escalated interest in orthopaedic surgery over the last two decades, as they have been demonstrated to ameliorate accuracy over conventional instruments. The next generation of computer assistance is being developed using Augmented Reality (AR). AR systems are broadly considered to exhibit some essential advantages over traditional computer navigation platforms, as they diminish the requirement for massive external detection equipment or reflecting markers by utilizing little detectors with built-in detecting mechanisms attached to the

apposite places in the operating field [3,4]. Also, AR systems compared to conventional computer navigation offer patient-specific CT imaging, thus developing a personalized initial plan for the surgeon, while also being able to visualize the mechanical axis and registration points. AR systems' hardware is featured as more compact and uncomplicated contrasted to computer navigation, which could lead to reduction in associated capital expenditures and maintenance costs [4,5]. Nonetheless, the exactness of augmented reality navigation has yet to be established [3]. These systems' utilization ranges from preoperatively surgical planning and simulation to navigation systems assisting surgeons intraoperatively [5]. With the implementation of navigation systems, orthopaedic surgeons can precisely track and visualize surgical instruments in real-time, conforming to the anatomical structures [3].

The growingly prompt evolution of AR technologies features the prospect of attaining the exemplary form of the human-machine interface [6]. Each AR system comprises distinct hardware and software, which provides the surgeon with real-time computer-processed imaging data. AR systems project their information to the surgeon in a way that combines real-life objects with superimposed computer-generated images [7-11]. A system control software utilizes the data from a position-tracking system, transforming the input into images, which are conveyed to a display system, where the amalgamation with the real-scene view transpires in front of the surgeons' eyes [12,13].

Knee+™ AR navigation system (Pixee medical company, Besancon, France) comprises a pair of smart glasses worn by the surgeon, specific markers (QR-code) connected to the tibial and femoral cutting guides and a central laptop (Figure 1). The Knee+™ system assists the surgeon in determining reference alignment axes in relation to anatomical landmarks for precise positioning of cutting guides regarding computed mechanical axis (Figure 2). The smart glasses (Vuzix 2000) enable the surgeon to visualize the tibial and femoral axis superimposed on the patient in real-time during operation, providing meaningful information concerning surgical decisions. With this information, surgeons can accurately specify the distal femoral cut, flexion/extension gaps, the varus/valgus axis, and the tibial cut's posterior slope. Smart glasses can also offer real-time

information to the surgeon regarding the accuracy of osteotomies according to the operative plan, combining the difference in the axis between the anatomical cut and the computerized plan. At this point, the system allows the surgeon to deliberate on the osteotomies and reach a new surgical plan.

This case series endeavors to scrutinize the accurateness of component alignment and restoration of the affected limb's mechanical axis in primary TKA, employing the Knee+™ system. Additionally, this study aims to evaluate whether such systems are substantively beneficial for a high-experienced knee surgeon.

MATERIALS AND METHODS

Study population

In our study, 30 patients underwent a primary unilateral TKA for osteoarthritis with AR guidance from May 2021 to December 2021. The average age of patients was 71.6 years, with five men and 25 women. Patients were included irrespective of age, diagnosis, deformity and body mass index (BMI). Revision surgery cases were excluded.

Preoperative examination

Preoperatively, radiographic exams (simple A/P and lateral knee x-rays) and full leg-length A/P x-rays were performed to all patients to assess the axial limb alignment (Figure 3A). The radiographic examinations were performed in the same center, utilizing the same software protocol. All radiological views were obtained with the patients on standing position. With regard to full leg-length A/P x-rays particulars, the images included in each set of radiographs were hip, knee and ankle views, with each image dimension being 3408*3320 pixels and with a resolution of 150*150, as well as a full-length lower limb x-ray performed by the same versed radiologist. The digital radiography system employed was the "NOVA FA" (Sedecal , Madrid, Spain), which featured a flat panel detector. The distance between the x-ray source and the patient was 160 cm. This specific system was a linear system in which the x-ray source moved from top to bottom to capture the images. It is pivotal to underline that this special

preoperative imaging examination is routinely not necessary for the utilization of Knee+™ system intraoperatively, and we carried out this protocol for the purposes of our study.

Surgery

All patients were operated on by the same adept orthopaedic surgeon, under combined regional anesthesia and without a tourniquet. The following prostheses were employed: Evolution® Medial-Pivot Knee System / MicroPort Orthopedics (7 cases), BalanSys BICONDYLAR® Knee System / Mathys European Orthopaedics (12 cases) and Vanguard® Complete Knee System / Zimmer Biomet (13 cases).

All operations were conducted with the assistance of the Knee+™ augmented reality navigation system (Pixee medical company, Besancon, France), using a standardized protocol of cutting first the tibial and then the femoral component. Concerning intraoperative soft tissue balance assessment, depending on the type of implant utilized, pertinent spacer blocks were employed for diligent evaluation of flexion and extension gaps in valgus and varus stress.

Postoperative course and imaging

Acute postoperatively, plain knee x-rays (A/P and lateral) were carried out. Subsequently, all patients were mobilized 2-3 h after the surgery with the aid of physiotherapists specialized in rapid recovery rehabilitation techniques. On the second postoperative day, following intensive physiotherapy sessions, a new full leg-length x-ray was performed in each patient, employing the same radiographic exam protocol as pre-surgery, for confirming the implants' position, orientation and alignment in the coronal plane (Figure 3B).

Data collection

We recorded measurements in three steps during the entire procedure for the femoral varus and flexion, for the tibial varus and posterior slope (Figure 4). At first, we documented the expected values preoperatively after the evaluation of joint deformity and the mechanical axis from the AR system. Afterwards, we recorded the same

measurements after each cut intraoperatively, and ultimately, we also measured these values radiologically after the operation.

Statistical analysis

Lin's concordance correlation coefficient (CCC) was estimated in terms of statistical analysis. Also, results for Bland and Altman's limits-of-agreement (LOA) procedure are provided as the mean of the two values, minus and plus 1.96 standard deviations. CCCs between 0.60 and 0.80 are considered substantial, while coefficients greater than 0.80 are considered excellent. As the discrepancies between the measurements could not be assumed to be normal, the Wilcoxon Signed Rank Test was also performed to examine whether there was a significant difference between the mean values of the expected values and the radiographic measures, as well as between the mean values of the expected and controlled values. If a statistically significant difference was detected, a Wilcoxon Signed Rank Test was carried out to test if the differences were significantly different from the 1 degree. The level of statistical significance was set to 0.05.

3

RESULTS

A total of 30 patients were included in the study. For patients' femur calculations, the difference between the controlled and the expected varus/valgus values ranged from -1 to 1 degree, whilst there was no discrepancy in terms of radiographic measurements. Regarding flexion values, the difference between expected and controlled values (which refer to the calculated varus/valgus values from the AR system after the osteotomies have been executed) ranged from -1 to 1 degree. The same was observed for the difference between expected values and radiographic measurements. The mean differences between all paired comparisons varied from 0 to 0.33 degrees (Table 1, 2).

Concerning tibia calculations, the discrepancy between controlled and expected values for varus ranged from -1 to 1 degree with a medium value of zero degrees, while the difference between radiographic measurements and expected values for varus ranged from 0 to 1 degree with a medium value of zero degrees. Finally, the difference between controlled and expected values for the posterior slope ranged from -2 to 1 degree and

between radiographic measurements and expected values from -1 to 1 (Table 1, 2). The corresponding median values were equal to zero. The mean differences between all paired comparisons were narrow, varying from 0 to 0.23 degrees (Table 3).

Near-perfect CCCs were reckoned for comparisons only between estimated flexion values and controlled and radiographic measurements in the femur and between estimated posterior slope values and controlled and radiographic measurements in the tibia, varying from 0.66 to 0.89. Also, as mentioned before, no deviation was observed between expected varus values and radiographic measurements in the femur. The 95% limits of agreement were within -1.46 to 1.52 degrees, and most estimates lie within the indicating (Table 3). Low CCC was estimated for expected and controlled values of varus in the femur and tibia. A graphical display of the data and their agreement are presented (Table 4).

A Wilcoxon Signed Rank Test was performed to determine whether there was a statistically significant difference regarding the expected values and radiographic calculations. The test revealed a statistically significant difference in mean expected values and radiographic measurements only for femoral flexion measurements (Z score = 2.67, p-value = 0.01). However, this difference was statistically significantly lower than 1 degree (Z score = -4.21, p-value < 0.01). Concerning discrepancies in the values of expected values and controlled measurements, a statistically significant difference between varus values measured in tibia was noted (Z score = -2.33, p-value = 0.02), which was also statistically significantly lower than 1 degree (Z score = -4.99, p-value < 0.01). Finally, it is of utmost importance to mention that there was no difference between the different implants used.

DISCUSSION

Our study aimed to denote the reproducible accuracy of an AR system (Knee+™, Pixee Medical) in 30 TKA patients compared to their intraoperative measurements. At this time, to the best of our knowledge, this is one of the largest clinical studies of this size to examine the accuracy and efficacy of this system, but the first to include implants with

different characteristics. The results indicate that good varus/valgus alignment was accomplished without outliers, whilst sagittal alignment was generally featured satisfactory. The procedure was performed comfortably and repeatedly by the same team. With regard to the rigorous limits and standards employed for the goals of our study, it is vitally important to emphasize that a difference of 1° in final implants' position is unlikely to be clinically significant, as numerous studies have demonstrated that differences in coronal alignment after TKA of up to 3° (or even up to roughly 6°) result in good clinical outcomes [14]. This exact point is crucial when considering the position of AR systems in the overall context of TKA.

There is an expanding interest in surgical variables intraoperatively controlled by orthopaedic surgeons, involving lower leg alignment, component positioning and soft-tissue balancing. Punctilious control over these factors is associated with improved outcomes, which is the main reason why several computer navigation and robotic-assisted systems have emerged [11-13].

In our study, three different types of implants were used with similar characteristics as surface arthroplasties. Their main difference lies in femoral flexion and posterior tibial slope. According to the aforementioned results, the Knee+™ AR system is a repeatable open platform method, eligible to use with the same accuracy regardless of the characteristics of the implants employed.

AR solutions can potentially decrease the outcome dependence on the surgeon's parameters by providing preoperative planning in the surgeon's field of view or even indicating impeccable trajectories for placing implants with overlays [11,12]. Diminishing cutting errors is the challenge of every AR system assisting total knee arthroplasty. According to literature data and the system's design, the system's accuracy lies principally in the varus/valgus tibial cut regarding the restoration of the mechanical axis.

More specifically, a recent study by Bennett KM *et al* suggested that coronal mean error for femoral and tibial cuts is 1.3° and 1.1°, respectively [13]. That present-day paper examined total knee arthroplasties carried out with the assistance of Knee+™ AR

system, as in our study. Comparing the two contemporary studies, results are featured comparatively similar, indicating that the robustness of this AR system regarding restoration of mechanical axis after primary total knee arthroplasty is perceptibly rigorous [13]. In a 2021 systematic review scrutinizing the accuracy of Knee+™ AR system by Iacono *et al*, it was deduced that this system is capable of executing cutting errors of less than 1° of discrepancy regarding coronal alignment and less than 2° in terms of posterior tibial slope and femoral extension/flexion, however, the copious limitations of this study need to be taken into consideration [15]. Perusing literature, it is patently evident that there are limited data concerning the effectiveness and accuracy of other AR systems used in TKA. Tsukada *et al* examined the efficacy of another AR system regarding the exactness of bone resection in TKA, concluding that AR technology can bolster surgeons' distal femoral resection accuracy compared to traditional intramedullary-guided techniques [16]. Additionally, another study by Fucentese *et al* examined the impact of "NextAR" AR TKA system (Medacta International SA, Castel San Pietro, Switzerland) in prosthesis alignment and positioning. It was inferred that, despite being employed in a low number of cases, the initial results seemed promising, requiring further research to corroborate the potency of this system [17].

The Knee+™ AR system equips the surgeon with a thorough preoperative plan projected in real-time concerning the alignment of bone components and mechanical axis restoration [18]. Notwithstanding, after initial cuts are executed, and the system describes a significant difference, it enables the surgeon to alter their surgical plans and correct the previously made osteotomies by contrasting the expected and controlled values. The design of this AR system does not require the utilization of intramedullary rods for femoral and tibial osteotomies, which are correlated with low-incidence intraoperative thromboembolic episodes [18].

All measurements are calculated intraoperatively, thus, there is no demand for time and cost-consuming preoperative radiographic scans that expose the patients to radiation [19,20]. During surgery, not using extra pins as markers for restoration of the mechanical

axis is linked with reduced risk of periprosthetic fractures during implantation, chiefly of the femoral component, and eliminated infection risk. Marker stabilization on the cutting guides is considered safe because it decreases the risk of systemic malalignment calculations due to intraoperative pin stretching, triggering imprecise cuts [21].

Although it has been demonstrated that AR might positively impact surgeons by diminishing surgical errors, verifying the learning curve would be an engrossing aim. Nonetheless, it is a user-friendly system with an effortless intraoperative setup that guides the surgeon throughout the surgery and is versatile in amending operative plans if necessary [22,23]. In our study, operative time and blood loss were not investigated, thus further research is incontrovertibly required to evaluate the potential impact of Knee+™ and other AR systems on this area.

Concerns have been raised regarding the extent of potentially spurious and disturbing AR information displayed intraoperatively. This AR system does not require substantial space or instrumentation in the operating room. Compared with other navigation systems, this digital tool allows the surgeon to concentrate on the surgical field through smart glasses and not get distracted by other display screens, allowing precious coordination between the surgical team. Apposite contrast and clarity of the AR technology and avoidance of masking structures in real patient view are of paramount significance [23].

CONCLUSION

Present-day literature data propound that AR systems, such as Knee+™, are becoming comparable to conventional navigation techniques in terms of precision and safety for routine clinical practice. AR appears to be a robust contemporary digital tool capable of revolutionizing the field of orthopaedic surgery, providing substantive information regarding intraoperative guidance and decision-making. In the future, it will distinctly possibly serve as a transcendent human-computer interface, enabling dexterous surgeons to attain superior results. Nonetheless, further technological and medical research is requisite to achieve augmented reality technologies' maximum potential and

cost-effectiveness. Until then, it is vitally important for the orthopaedic surgeon to rely on their training and adroitness for decision-making and opting wisely to employ an AR system.

ARTICLE HIGHLIGHTS

Research background

Computer-assisted systems obtained an increased interest in orthopaedic surgery over the last years, as they enhance precision compared to conventional hardware. The expansion of computer assistance is evolving with the employment of ² augmented reality. Yet, the accuracy of augmented reality navigation systems has not been determined.

Research motivation

The results indicate satisfactory postoperative coronal alignment without outliers across all three different implants utilized. Augmented reality navigation systems can bolster orthopaedic surgeons' accuracy in achieving precise axial alignment. However, further research is required to further evaluate their efficacy and potential.

Research objectives

This study aims to examine the accuracy of component alignment and restoration of the affected limb's mechanical axis in primary TKA, utilizing an augmented reality navigation system and to assess whether such systems are conspicuously fruitful for an accomplished knee surgeon.

Research methods

From May 2021 to December 2021, 30 patients, 25 women and five men, underwent a primary unilateral total knee arthroplasty. Revision cases were excluded. A preoperative radiographic procedure was performed to evaluate the limb's axial alignment. All patients were operated on by the same team, without a tourniquet,

utilizing three distinct prostheses with the assistance of the Knee+™ augmented reality navigation system in every operation. Postoperatively, the same radiographic exam protocol was executed to evaluate the implants' position, orientation and coronal plane alignment. We recorded measurements in 3 stages regarding femoral varus and flexion, tibial varus and posterior slope. Firstly, the expected values from the AR system were documented. Then we calculated the same values after each cut and finally, the same measurements were recorded radiologically after the operations. Concerning statistical analysis, Lin's concordance correlation coefficient was estimated, while Wilcoxon Signed Rank Test was performed when needed.

Research results

Further research is required to evaluate their efficacy and potential.

Research conclusions

The results indicate satisfactory postoperative coronal alignment without outliers across all three implants. Augmented reality navigation systems can bolster orthopaedic surgeons' accuracy in achieving precise axial alignment.

Research perspectives

A statistically significant difference was observed regarding mean expected values and radiographic measurements for femoral flexion measurements only (Z score = 2.67, p-value = 0.01). Nonetheless, this difference was statistically significantly lower than 1 degree (Z score = -4.21, p-value < 0.01). In terms of discrepancies in the calculations of expected values and controlled measurements, a statistically significant difference between tibial varus values was detected (Z score = -2.33, p-value = 0.02), which was also statistically significantly lower than 1 degree (Z score = -4.99, p-value < 0.01).

5%

SIMILARITY INDEX

PRIMARY SOURCES

1	www.frontiersin.org Internet	89 words — 2%
2	www.ncbi.nlm.nih.gov Internet	79 words — 2%
3	D Gomez. "Anatomical distribution of colorectal cancer over a 10 year period in a district general hospital: is there a true "rightward shift"?", Postgraduate Medical Journal, 2004 Crossref	21 words — 1%

EXCLUDE QUOTES ON
EXCLUDE BIBLIOGRAPHY ON

EXCLUDE SOURCES < 1%
EXCLUDE MATCHES < 10 WORDS