



## Baishideng Publishing Group Co., Limited

Flat C, 23/F., Lucky Plaza,  
315-321 Lockhart Road,  
Wan Chai, Hong Kong, China

---

### ESPS Peer-review Report

**Name of Journal:** World Journal of Cardiology

**Ms:** 2853

**Title:** Longitudinal Stent Compression of Everolimus-eluting Stent: A Report of 2 Cases

**Reviewer code:** 00742205

**Science editor:** s.x.gou@wjgnet.com

**Date sent for review:** 2013-03-20 12:07

**Date reviewed:** 2013-03-21 00:17

| CLASSIFICATION                                     | LANGUAGE EVALUATION   | RECOMMENDATION                      | CONCLUSION   |
|--|---|-------------------------------------|--|
| <input type="checkbox"/> Grade A (Excellent)       | <input type="checkbox"/> Grade A: Priority Publishing                 | Google Search:                      | <input type="checkbox"/> Accept                        |
| <input type="checkbox"/> Grade B (Very good)       | <input checked="" type="checkbox"/> Grade B: minor language polishing | <input type="checkbox"/> Existed    | <input type="checkbox"/> High priority for publication |
| <input checked="" type="checkbox"/> Grade C (Good) | <input type="checkbox"/> Grade C: a great deal of language polishing  | <input type="checkbox"/> No records | <input type="checkbox"/> Rejection                     |
| <input type="checkbox"/> Grade D (Fair)            |   | BPG Search:                         | <input type="checkbox"/> Minor revision                |
| <input type="checkbox"/> Grade E (Poor)            | <input type="checkbox"/> Grade D: rejected                            | <input type="checkbox"/> Existed    | <input type="checkbox"/> Major revision                |
|  |   | <input type="checkbox"/> No records |  |

### COMMENTS

#### COMMENTS TO AUTHORS:

I feel the cases represent important experience to all the cardiologists who perform similar procedure. But the manuscript can be shortened significantly. Similar type of catheter was described multiple times in on paragraph becomes very redundant.



## Baishideng Publishing Group Co., Limited

Flat C, 23/F., Lucky Plaza,  
315-321 Lockhart Road,  
Wan Chai, Hong Kong, China

---

### ESPS Peer-review Report

**Name of Journal:** World Journal of Cardiology

**Ms:** 2853

**Title:** Longitudinal Stent Compression of Everolimus-eluting Stent: A Report of 2 Cases

**Reviewer code:** 02457603

**Science editor:** s.x.gou@wjgnet.com

**Date sent for review:** 2013-03-20 12:07

**Date reviewed:** 2013-03-29 07:48

| CLASSIFICATION                                     | LANGUAGE EVALUATION   | RECOMMENDATION                      | CONCLUSION   |
|--|---|-------------------------------------|--|
| <input type="checkbox"/> Grade A (Excellent)       | <input type="checkbox"/> Grade A: Priority Publishing                 | Google Search:                      | <input type="checkbox"/> Accept                        |
| <input type="checkbox"/> Grade B (Very good)       | <input checked="" type="checkbox"/> Grade B: minor language polishing | <input type="checkbox"/> Existed    | <input type="checkbox"/> High priority for publication |
| <input type="checkbox"/> Grade C (Good)            | <input type="checkbox"/> Grade C: a great deal of                     | <input type="checkbox"/> No records | <input checked="" type="checkbox"/> Rejection          |
| <input checked="" type="checkbox"/> Grade D (Fair) | language polishing  | BPG Search:                         | <input type="checkbox"/> Minor revision                |
| <input type="checkbox"/> Grade E (Poor)            | <input type="checkbox"/> Grade D: rejected                            | <input type="checkbox"/> Existed    | <input type="checkbox"/> Major revision                |
|  |   | <input type="checkbox"/> No records |  |

### COMMENTS

#### COMMENTS TO AUTHORS:

If I were you, I would try to do the LAD again before I handle the RCA or advise the patient to accept CABG in case 2. Which will be more safer for the patient.



## Baishideng Publishing Group Co., Limited

Flat C, 23/F., Lucky Plaza,  
315-321 Lockhart Road,  
Wan Chai, Hong Kong, China

### ESPS Peer-review Report

**Name of Journal:** World Journal of Cardiology

**Ms:** 2853

**Title:** Longitudinal Stent Compression of Everolimus-eluting Stent: A Report of 2 Cases

**Reviewer code:** 00060499

**Science editor:** s.x.gou@wjgnet.com

**Date sent for review:** 2013-03-20 12:07

**Date reviewed:** 2013-04-04 00:19

| CLASSIFICATION           | LANGUAGE EVALUATION                             | RECOMMENDATION | CONCLUSION                        |
|--------------------------|---|----------------|-----------------------------------|
| [ ] Grade A (Excellent)  | [ Y] Grade A: Priority Publishing               | Google Search: | [ ] Accept                        |
| [ Y] Grade B (Very good) | [ ] Grade B: minor language polishing           | [ ] Existed    | [ ] High priority for publication |
| [ ] Grade C (Good)       | [ ] Grade C: a great deal of language polishing | [ ] No records | [ ] Rejection                     |
| [ ] Grade D (Fair)       | [ ] Grade D: rejected                           | BPG Search:    | [ ] Minor revision                |
| [ ] Grade E (Poor)       |   | [ ] Existed    | [ Y] Major revision               |
|                          |   | [ ] No records |                                   |

### COMMENTS

#### COMMENTS TO AUTHORS:

-Well written paper -Some confusion in the naming of figures -was stent boost used?? -Need to describe the mechanisms more appropriately in addition to guide catheter induced LSC -Need to describe the details of types of stents available, design and structure, radial and longitudinal strength of stents, relation between newer materials used for stents and their strength, reasoning for their use, advantages and disadvantages of these newer material w.r.t alloy used, strut thickness, radial and longitudinal strength compared to stainless steel stents, is LSC class effect? -no mention of 'concertina effect' - is there relation between thinness of struts, number of connectors between struts, and the orientation of those connectors, -Need to provide some more advise to interventionists regarding lesion characteristics and which stent to select as well as preparation of lesion to prevent this effect, post stenting advise whether compliant or NC balloon and why?.